



MS ISO/IEC 17025

# *Certificate of Accreditation*

No: SMM 122

Valid until: 21 October 2011

This is to certify that

**INTEGRAJAYA CALIBRATION TECHNOLOGIES SDN BHD**  
**PULAU PINANG**  
**MALAYSIA**  
(FIELDS OF CALIBRATION: ELECTRICAL & TEMPERATURE)

has been granted accreditation in respect of the scope of accreditation described in the SCHEDULE attached, subject to the terms and conditions governing the *Skim Akreditasi Makmal Malaysia (SMM)*, the Laboratory Accreditation Scheme of Malaysia.

Laboratories accredited under SMM meet the requirements of MS ISO/IEC 17025 'General requirements for the competence of testing and calibration laboratories'. This Malaysian Standard is identical with ISO/IEC 17025 published by the International Organization for Standardization (ISO).

*"This laboratory is accredited in accordance with recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communiqué dated 18 June 2005)"*



**(RIDZWAN KASIM)**

for the Director General  
Department of Standards Malaysia  
Date of issue: 20 October 2008

# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

## NO. SMM 122

(Issue 3, 26 October 2010 replacement of SMM 122 dated 5 October 2009)

Page 1 of 15

**LABORATORY LOCATION:**  
(PERMANENT LABORATORY)

**INTEGRAJAYA CALIBRATION TECHNOLOGIES SDN BHD**  
**NO. 737-5-5, KOMPLEKS SRI SUNGAI NIBONG**  
**JALAN SULTAN AZLAN SHAH**  
**11900 PENANG, MALAYSIA**

The standard used for assessment of this laboratory is MS ISO/IEC 17025:2005

\* The expanded uncertainties are based on an estimated confidence probability of not less than 95% and have a coverage factor of k=2 unless stated otherwise.

<sup>(1)</sup> The CMC is expressed as  $\pm$  (Of Indication in  $\mu\text{V/V}$  + floor value in  $\mu\text{V}$ )

**FIELD OF CALIBRATION**

**ELECTRICAL**

**SCOPE OF ACCREDITATION:**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks
<b>1. Indicating instrument</b>			
DC VOLTAGE	0 $\mu\text{V}$ to 200 $\mu\text{V}$ 0.2 mV to 2.0 mV 2 mV to 20 mV 20 mV to 200 mV 0.2 V to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	6.4 $\mu\text{V/V}$ + 0.44 $\mu\text{V}$ 6.5 $\mu\text{V/V}$ + 0.43 $\mu\text{V}$ 6.9 $\mu\text{V/V}$ + 0.43 $\mu\text{V}$ 7.9 $\mu\text{V/V}$ + 0.42 $\mu\text{V}$ 6.5 $\mu\text{V/V}$ + 0.82 $\mu\text{V}$ 6.0 $\mu\text{V/V}$ + 3.3 $\mu\text{V}$ 9.5 $\mu\text{V/V}$ + 55 $\mu\text{V}$ 8.7 $\mu\text{V/V}$ + 0.41 mV	Generation using calibrator model Wavetek 4800A
DC CURRENT	0 to 200 $\mu\text{A}$ 0.2 m to 2 mA 2 mA to 20 mA 20 mA to 200 mA 0.2 A to 2 A	97 $\mu\text{A/A}$ + 1.7 nA 51 $\mu\text{A/A}$ + 17 nA 57 $\mu\text{A/A}$ + 85 nA 50 $\mu\text{A/A}$ + 0.84 $\mu\text{A}$ 0.11 mA/A + 25 $\mu\text{A}$	Generation using calibrator model Wavetek 4800A
	0.32 A to 3.2 A 3.2 A to 10.5 A	0.68 mA/A + 0.14 mA 0.61 mA/A + 1.1 mA	Generation using calibrator model Wavetek 9100



# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

## NO. SMM 122

(Issue 3, 26 October 2010 replacement of  
SMM 122 dated 5 October 2009)

Page 2 of 15

### FIELD OF CALIBRATION

### ELECTRICAL

### SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks	
<b>1. Indicating instrument</b>				
DC CURRENT	3.2 A to 32 A 32 A to 105 A 105 A to 200 A	1.3mA/A + 50mA 2.2mA/A + 25mA 2.3mA/A + 29mA	Generation using calibrator model Wavetek 9100 & 10 turn coil	
	16 A to 160 A 160 A to 525 A 525 A to 1000 A	2.2mA/A + 29mA 2.3mA/A + 23mA 2.3mA/A + 0.1A	Generation using calibrator model Wavetek 9100 & 50 turn coil	
AC VOLTAGE	0.09 mV to 1000 V	(See Matrix A)	Generation using calibrator model Wavetek 4800A	
AC CURRENT	0 $\mu$ A to 10.5 A	(See Matrix B)	Generation using calibrator model Wavetek 9100	
	<b>Range</b>	<b>Frequency</b>		
		<b>(0.01 to 1) kHz</b>	<b>(1 to 5) kHz</b>	Generation using calibrator model Wavetek 4800A
	9 $\mu$ A to 200 $\mu$ A	0.16 mA/A + 8.2 nA	0.34 mA/A + 12 nA	
	0.2 mA to 2 mA	0.12 mA/A + 82 nA	0.25 mA/A + 80 nA	
	2 mA to 20 mA	0.11 mA/A + 0.8 $\mu$ A	0.25 mA/A + 0.81 $\mu$ A	
	20 mA to 200 mA	0.11 mA/A + 8 $\mu$ A	0.25 mA/A + 8.1 $\mu$ A	
	0.2 A to 2 A	0.28 mA/A + 85 $\mu$ A	0.46 mA/A + 110 $\mu$ A	
	<b>Range</b>	<b>Frequency</b>		Generation using calibrator model Wavetek 9100 & 10/50 turn coil
		<b>(10 to 100) Hz</b>	<b>(100 to 440) Hz</b>	
3.2 A to 32 A	2.2 mA/A + 48 mA	8.5 mA/A + 52 mA		
32 A to 200 A	3.2 mA/A + 96 mA	7.9 mA/A + 280 mA		
16 A to 160 A	3.1 mA/A + 43 mA	-		
160 A to 1000 A	3.2 mA/A + 440 mA	-		
DC RESISTANCE (Specific Value)	10 $\Omega$	0.32 m $\Omega$	Generation using calibrator model Wavetek 4800A	
	100 $\Omega$	15 m $\Omega$		
	1 k $\Omega$	14 m $\Omega$		
	10 k $\Omega$	14 $\Omega$		
	100 k $\Omega$	18 $\Omega$		
	1 M $\Omega$	36 $\Omega$		
	10 M $\Omega$	0.72 k $\Omega$		
	100 M $\Omega$	7 k $\Omega$		



# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

## NO. SAMM 122

(Issue 3, 26 October 2010 replacement of SAMM 122 dated 5 October 2009)

Page 3 of 15

### FIELD OF CALIBRATION

### ELECTRICAL

### SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks	
<b>1. Indicating instrument</b>				
DC RESISTANCE	0 $\Omega$ to 40 $\Omega$	0.56 m $\Omega$ / $\Omega$ + 23 m $\Omega$	Generation using calibrator model Wavetek 9100	
	40 $\Omega$ to 400 $\Omega$	0.17 m $\Omega$ / $\Omega$ + 25 m $\Omega$		
	400 $\Omega$ to 4 k $\Omega$	0.17 m $\Omega$ / $\Omega$ + 0.1 $\Omega$		
	4 k $\Omega$ to 40 k $\Omega$	0.17 m $\Omega$ / $\Omega$ + 1 $\Omega$		
	40 k $\Omega$ to 400 k $\Omega$	0.20 m $\Omega$ / $\Omega$ + 12 $\Omega$		
	400 k $\Omega$ to 4 M $\Omega$	0.22 m $\Omega$ / $\Omega$ + 0.14 k $\Omega$		
	4 M $\Omega$ to 40 M $\Omega$	0.56 m $\Omega$ / $\Omega$ + 2.4 k $\Omega$		
	40 M $\Omega$ to 400 M $\Omega$	0.64 m $\Omega$ / $\Omega$ + 64 k $\Omega$		
CAPACITANCE	Range	Frequency		Generation using calibrator model Wavetek 9100
		$\leq$ 350 Hz	0.35 to 1.5 kHz	
	0.5 nF to 4.0 nF	3.4 + 17 pF	6.8 + 34 pF	
	4 nF to 40 nF	3.4 + 34 pF	6.8 + 68 pF	
	40 nF to 400 nF	3.4 + 0.18 nF	6.8 + 0.36 nF	
	400 nF to 4 $\mu$ F	4.5 + 1.8 nF	9.1 + 3.6 nF	
	4 $\mu$ F to 40 $\mu$ F	5.7 + 18 nF	11 + 36 nF	
	40 $\mu$ F to 400 $\mu$ F	5.7 + 0.18 $\mu$ F	11 + 0.36 $\mu$ F	
400 $\mu$ F to 4 mF	5.7 + 1.8 $\mu$ F	11 + 3.6 $\mu$ F		
4 mF to 40 mF	11 + 68 $\mu$ F	23 + 0.14 mF		
CAPACITANCE (Specific Value)	1 kHz 1 pF 10 pF 100 pF 1000 pF	1.2 fF 12 fF 0.12 pF 1.2 pF	Generation using calibrator model Agilent 16380A	
	1 kHz 0.01 $\mu$ F 0.1 $\mu$ F 1 $\mu$ F	11 pF 0.11 nF 1.1 nF	Generation using calibrator model Agilent 16380C	



**NO. SAMM 122**

(Issue 3, 26 October 2010 replacement of SAMM 122 dated 5 October 2009)

**FIELD OF CALIBRATION**

**ELECTRICAL**

**SCOPE OF ACCREDITATION:**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks
<b>1. Indicating instrument</b>			
INDUCTANCE	1 kHz 1mH to 900 mH	23 mH/H + 0.57 $\mu$ H	Generation using calibrator model IET LS-400A
INDUCTANCE (Specific Value)	100 Hz, 120Hz, 1 kHz, 10 kHz, 100 kHz 100 $\mu$ H 1 mH 10 mH 100 mH 1 H	61 nH 0.61 $\mu$ H 6.1 $\mu$ H 61 $\mu$ H 0.61 mH	Generation using calibrator model General Radio 1482
FREQUENCY	0.1 Hz to 1 GHz	85 nHz/Hz	Using 10 MHz Reference model Fluke PM6680B
<b>2. Oscilloscope</b>			
a) Bandwidth	100 mHz to 100 MHz 100.01 MHz to 550 MHz 550.01 MHz to 1.1 GHz	22 mHz/Hz 34 mHz/Hz 43 mHz/Hz	Generation using calibrator model Fluke 9500/9510
b) Time Base	450 ps to 900 ps @ 100 mV to 500 mV  900 ps to 55 sec @ 100 mV to 1 V	1.1 ms/s  1.1 ms/s	Generation using calibrator model Fluke 9500/9510



# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

## NO. SMM 122

(Issue 3, 26 October 2010 replacement of SMM 122 dated 5 October 2009)

### FIELD OF CALIBRATION

### ELECTRICAL

### SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*				Remarks
<b>2. Oscilloscope</b>						
c) Amplitude	Range	DC Signal ( $\pm$ )		Square Signal (10Hz to 10kHz)		Generation using calibrator model Fluke 9500/9510
		Into 50 $\Omega$	Into 1M $\Omega$	Into 50 $\Omega$	Into 1M $\Omega$	
	35 $\mu$ V to 1 mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 25 $\mu$ V	10mV/V + 1.1 $\mu$ V	10mV/V + 1.1 $\mu$ V	
	1mV to 21mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 25 $\mu$ V	2.4mV/V + 3.6 $\mu$ V	2.5mV/V + 1 $\mu$ V	
	21mV to 100mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 25 $\mu$ V	0.99mV/V + 1.7 $\mu$ V	0.99mV/V + 1.8 $\mu$ V	
	100mV to 556mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 45 $\mu$ V	0.98mV/V + 16 $\mu$ V	0.98mV/V + 16 $\mu$ V	
	556mV to 5.5V	0.25mV/V + 26 $\mu$ V	0.25mV/V + 45 $\mu$ V	0.44mV/V + 0.33mV	0.47mV/V + 0.3mV	
	5.5V to 10V	-	0.25mV/V + 45 $\mu$ V	-	0.47mV/V + 0.3mV	
	10V to 100V	-	0.25mV/V + 81 $\mu$ V	-	0.40mV/V + 35mV	
100V to 210V	-	0.25mV/V + 93 $\mu$ V	-	0.40mV/V + 35mV		
210V to 222V	-	0.25mV/V + 93 $\mu$ V	-	-		
<b>3. Source</b>						
DC VOLTAGE	0 to 200 mV 0.2 V to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	9.1 $\mu$ V/V + 0.11 $\mu$ V 7.6 $\mu$ V/V + 0.41 $\mu$ V 7.2 $\mu$ V/V + 2.1 $\mu$ V 12 $\mu$ V/V + 43 $\mu$ V 12 $\mu$ V/V + 42 $\mu$ V		Measurement using Multimeter model Wavetek 1281		
	1 kV to 10 kV	0.57 mV/V + 0.57 V		Measurement using High Voltage Meter model Vallaha 4600		
AC VOLTAGE	0 to 1000 V	(See Matrix C)		Measurement using Multimeter model Wavetek 1281		
	1 kV to 2 kV @ 20 Hz to 60 Hz	1.1 mV/V + 2.3 V		Measurement using High Voltage Meter model Vallaha 4600		
	2 kV to 15 kV @ 20 Hz to 60 Hz	5.7 mV/V + 57 V		Measurement using High Voltage Meter model Vallaha 4600		
DC CURRENT	0 to 200 $\mu$ A 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 0.2 A to 2 A 2 A to 10 A	0.12 mA/A + 0.45 nA 0.12 mA/A + 4.5 nA 0.12 mA/A + 45 nA 0.11 mA/A + 1.1 $\mu$ A 0.23 mA/A + 22 $\mu$ A 0.10 mA/A + 0.6 $\mu$ A		Measurement using Multimeter model Wavetek 1281		



# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

## NO. SAMM 122

(Issue 3, 26 October 2010 replacement of SAMM 122 dated 5 October 2009)

Page 6 of 15

### FIELD OF CALIBRATION

### ELECTRICAL

### SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks			
<b>3. Source</b>						
AC CURRENT	0 to 2 A	(See Matrix D)	Measurement using Multimeter model Wavetek 1281			
	2 A to 10 A	(See Matrix D)	Measurement using Multimeter & Current Shunt model Wavetek 1281 & Wavetek 4953			
DC RESISTANCE	0 to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 0.2 k $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 0.2 M $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$ 0.2 G $\Omega$ to 1G $\Omega$	18 $\mu\Omega/\Omega$ + 22 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 65 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 0.63 m $\Omega$ 11 $\mu\Omega/\Omega$ + 6.3 m $\Omega$ 13 $\mu\Omega/\Omega$ + 56 m $\Omega$ 20 $\mu\Omega/\Omega$ + 1.3 $\Omega$ 37 $\mu\Omega/\Omega$ + 87 $\Omega$ 0.35 m $\Omega/\Omega$ + 10 k $\Omega$ 3.5 m $\Omega/\Omega$ + 0.5 M $\Omega$	Measurement using Multimeter model Wavetek 1281			
RF POWER	Range -30 dBm to 20 dBm (1 $\mu$ W to 100 mW into 50 $\Omega$ )	Frequency				Measurement using Power Meter & Power Sensor model Agilent EPM441A & 8482A
		(100 to 300) kHz	(0.3 to 1) MHz	(0.001 to 2) GHz	(2.0 to 4.2) GHz	
		2.8 %	2.3 %	2.3 %	2.4 %	
FREQUENCY	100 mHz to 2.7 GHz	85 nHz/Hz		Measurement using Frequency Counter model Fluke PM6680B		

### Signatory:

NO. NAME

I/C NO.

1. Lam Sik Lee

731122-07-5254



**NO. SAMM 122**(Issue 3, 26 October 2010 replacement of  
SAMM 122 dated 5 October 2009)

Page 7 of 15

**FIELD OF CALIBRATION****ELECTRICAL****SCOPE OF ACCREDITATION:****Matrix A - AC Voltage Measurement**

Range	Frequency						
	(10 to 31) Hz	(32 to 330) Hz	(0.3 to 10) kHz	(10 to 33) kHz	(30 to 100) kHz	(100 to 330) kHz	(0.300 to 1) MHz
0.09 mV to 2 mV	0.20 + 0.0051	0.20 + 0.0051	0.17 + 0.0051	0.27 + 0.0051	0.52 + 0.0051	1.1 + 0.021	2.2 + 0.024
0.2 mV to 20 mV	0.18 + 0.005	0.15 + 0.005	0.15 + 0.005	0.23 + 0.005	0.55 + 0.0049	1.1 + 0.021	2.3 + 0.024
20 mV to 200 mV	0.17 + 0.0077	0.13 + 0.0077	0.11 + 0.0074	0.21 + 0.0071	0.44 + 0.007	1.2 + 0.033	2.3 + 0.12
0.2 V to 2 V	0.12 + 0.026	0.073 + 0.017	0.074 + 0.0083	0.074 + 0.0082	0.13 + 0.017	0.39 + 0.084	2.1 + 0.34
2 V to 20 V	0.12 + 0.25	0.073 + 0.17	0.074 + 0.083	0.074 + 0.082	0.14 + 0.17	0.36 + 0.84	2.3 + 4.2
20 V to 200 V	0.17 + 2.6	0.11 + 1.7	0.074 + 0.82	0.083 + 1.7	0.27 + 2.5	0.77 + 42	8.2 + 110
200 V to 1000 V	0.2 + 17	0.2 + 17	0.16 + 17	0.19 + 33	0.28 + 10	-	-

The expanded uncertainties given in this table are expressed in **mV/V + mV****Matrix B - AC Current Measurement**

Range	Frequency			
	(0.01 to 3) kHz	(3 to 10) kHz	(10 to 20) kHz	(20 to 30) kHz
0 $\mu$ A to 32 $\mu$ A	0.79 + 1	1.1 + 2	2.3 + 6.8	2.8 + 10
32 $\mu$ A to 320 $\mu$ A	0.79 + 0.35	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
0.32 mA to 3.2 mA	0.75 + 0.48	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
3.2 mA to 32 mA	0.79 + 4.7	1.1 + 7.3	2.3 + 15	2.8 + 25
32 mA to 320 mA	0.91 + 36	1.1 + 55	2.3 + 73	2.8 + 110
0.32 A to 3.2 A	1.1 + 570	2.8 + 2900	-	-
3.2 A to 10.5 A	2.3 + 3400	5.7 + 11000	-	-

The expanded uncertainties given in this table are expressed in **mA/A +  $\mu$ A**



**NO. SAMM 122**(Issue 3, 26 October 2010 replacement of  
SAMM 122 dated 5 October 2009)

Page 8 of 15

**FIELD OF CALIBRATION****ELECTRICAL****SCOPE OF ACCREDITATION:****Matrix C - AC Voltage Source**

Range	Frequency							
	(10 – 40) Hz	(40 – 100) Hz	(0.1 – 2) kHz	(2 – 10) kHz	(10 – 30) kHz	(30 – 100) kHz	(100 – 300) kHz	(0.3 – 1) MHz
0 to 200mV	0.27 + 0.0044	0.25 + 0.0045	0.23 + 0.0023	0.23 + 0.0045	0.48 + 0.0089	0.90 + 0.022	-	-
0.2V to 2V	0.20 + 0.022	0.17 + 0.022	0.15 + 0.022	0.17 + 0.022	0.28 + 0.045	0.57 + 0.23	3.4 + 2.3	11 + 23
2V to 20V	0.20 + 0.22	0.17 + 0.22	0.15 + 0.22	0.17 + 0.22	0.28 + 0.45	0.57 + 2.3	3.4 + 2.3	11 + 230
20V to 200V	0.20 + 2.2	0.17 + 2.2	0.15 + 2.2	0.17 + 2.2	0.28 + 4.5	0.57 + 23	3.4 + 230	11 + 2300
200V to 1000V	0.22 + 10	0.20 + 10	0.20 + 10	0.20 + 10	0.30 + 22	0.60 + 110	-	-

The expanded uncertainties given in this table are expressed in **mV/V + mV****Matrix D - AC Current Source**

Range	Frequency				
	(10 – 40) Hz	(40 – 300) Hz	(0.3 – 1) kHz	(1 – 5) kHz	(5 – 10) kHz
0 to 200 $\mu$ A	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	-
0.2 mA to 2 mA	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	-
2 mA to 20 mA	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	-
20 mA to 200 mA	0.36 + 22	0.36 + 22	0.36 + 22	0.36 + 22	-
0.2 A to 2 A	0.69 + 450	0.69 + 450	0.69 + 450	2.3 + 910	-
2 A to 10 A	0.63 + 110	0.61 + 100	0.61 + 48	0.61 + 100	0.61 + 100

The expanded uncertainties given in this table are expressed in **mA/A +  $\mu$ A**

**NO. SAMM 122**

(Issue 3, 26 October 2010 replacement of SAMM 122 dated 5 October 2009)

**FIELD OF CALIBRATION**

**ELECTRICAL**

**SITE CALIBRATION – CATEGORY 1**

**SCOPE OF ACCREDITATION:**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks	
<b>1. Indicating instrument</b>				
DC VOLTAGE	0 $\mu$ V to 200 $\mu$ V 0.2 mV to 2.0 mV 2 mV to 20 mV 20 mV to 200 mV 0.2 V to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	6.4 $\mu$ V/V + 0.44 $\mu$ V 6.5 $\mu$ V/V + 0.43 $\mu$ V 6.9 $\mu$ V/V + 0.43 $\mu$ V 7.9 $\mu$ V/V + 0.40 $\mu$ V 6.6 $\mu$ V/V + 0.82 $\mu$ V 6.0 $\mu$ V/V + 3.3 $\mu$ V 9.5 $\mu$ V/V + 55 $\mu$ V 8.7 $\mu$ V/V + 0.41 mV	Generation using calibrator model Wavetek 4800A	
DC CURRENT	0 to 200 $\mu$ A 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 0.2 A to 2 A	97 $\mu$ A/A + 1.7 nA 51 $\mu$ A/A + 17 nA 57 $\mu$ A/A + 85 nA 50 $\mu$ A/A + 0.84 $\mu$ A 0.11 mA/A + 25 $\mu$ A	Generation using calibrator model Wavetek 4800A	
	0.32 A to 3.2 A 3.2 A to 10.5 A	0.6 mA/A + 0.44 mA 0.6 mA/A + 1.2 mA	Generation using calibrator model Wavetek 9100	
AC VOLTAGE	0.09 mV to 1000 V	(See Matrix E)	Generation using calibrator model Wavetek 4800A	
AC CURRENT	0 $\mu$ A to 10.5 A	(See Matrix F)	Generation using calibrator model Wavetek 9100	
	Range	Frequency		Generation using calibrator model Wavetek 4800A
			(0.01 to 1) kHz	
	9 $\mu$ A to 200 $\mu$ A	0.16 mA/A + 8.2 nA	0.34 mA/A + 12 nA	
	0.2 mA to 2 mA	0.12 mA/A + 82 nA	0.25 mA/A + 80 nA	
	2 mA to 20 mA	0.11 mA/A + 0.8 $\mu$ A	0.25 mA/A + 0.31 $\mu$ A	
20 mA to 200 mA	0.11 mA/A + 8 $\mu$ A	0.25 mA/A + 8.1 $\mu$ A		
0.2 A to 2 A	0.28 mA/A + 83 $\mu$ A	0.46 mA/A + 110 $\mu$ A		



**NO. SMM 122**

(Issue 3, 26 October 2010 replacement of SMM 122 dated 5 October 2009)

**FIELD OF CALIBRATION**

**ELECTRICAL**

**SITE CALIBRATION – CATEGORY 1**

**SCOPE OF ACCREDITATION:**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks	
<b>1. Indicating instrument</b>				
DC RESISTANCE (Specific Value)	10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$ 100 M $\Omega$	0.32 m $\Omega$ 15 m $\Omega$ 14 m $\Omega$ 14 $\Omega$ 18 $\Omega$ 36 $\Omega$ 0.72 k $\Omega$ 7 k $\Omega$	Generation using calibrator model Wavetek 4800A	
DC RESISTANCE	0 $\Omega$ to 40 $\Omega$ 40 $\Omega$ to 400 $\Omega$ 400 $\Omega$ to 4 k $\Omega$ 4 k $\Omega$ to 40 k $\Omega$ 40 k $\Omega$ to 400 k $\Omega$ 400 k $\Omega$ to 4 M $\Omega$ 4 M $\Omega$ to 40 M $\Omega$ 40 M $\Omega$ to 400 M $\Omega$	0.56 m $\Omega/\Omega$ + 23 m $\Omega$ 0.17 m $\Omega/\Omega$ + 25 m $\Omega$ 0.17 m $\Omega/\Omega$ + 0.1 $\Omega$ 0.17 m $\Omega/\Omega$ + 1 $\Omega$ 0.20 m $\Omega/\Omega$ + 12 $\Omega$ 0.22 m $\Omega/\Omega$ + 0.14 k $\Omega$ 0.56 m $\Omega/\Omega$ + 2.4 k $\Omega$ 0.64 m $\Omega/\Omega$ + 64 k $\Omega$	Generation using calibrator model Wavetek 9100	
CAPACITANCE	Range	Frequency		Generation using calibrator model Wavetek 9100
		$\leq$ 350 Hz	0.35 to 1.5 kHz	
	0.5 nF to 4.0 nF	3.4 + 17 pF	6.8 + 34 pF	
	4 nF to 40 nF	3.4 + 34 pF	6.8 + 68 pF	
	40 nF to 400 nF	3.4 + 0.18 nF	6.8 + 0.36 nF	
	400 nF to 4 $\mu$ F	4.5 + 1.8 nF	9.1 + 3.6 nF	
	4 $\mu$ F to 40 $\mu$ F	5.7 + 18 nF	11 + 36 nF	
	40 $\mu$ F to 400 $\mu$ F	5.7 + 0.18 $\mu$ F	11 + 0.36 $\mu$ F	
400 $\mu$ F to 4 mF	5.7 + 1.8 $\mu$ F	11 + 3.6 $\mu$ F		
4 mF to 40 mF	11 + 68 $\mu$ F	23 + 0.14 mF		
FREQUENCY	0.1 Hz to 1 GHz	85 nHz/Hz	Using 10 MHz Reference model Fluke PM6680B	



# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

## NO. SAMP 122

(Issue 3, 26 October 2010 replacement of  
SAMP 122 dated 5 October 2009)

Page 11 of 15

FIELD OF CALIBRATION

ELECTRICAL

SITE CALIBRATION – CATEGORY 1

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*				Remarks
<b>2. Oscilloscope</b>						
a. Bandwidth	100 MHz to 100 MHz 100.01 MHz to 550 MHz 550.01 MHz to 1.1 GHz	22 mHz/Hz 34 mHz/Hz 43 mHz/Hz		Generation using calibrator model Fluke 9500/9510		
b. Time Markers	450 ps to 900 ps 100 mV to 500 mV  900 ps to 55 sec 100 mV to 1 V	1.1 ms/s  1.1 ms/s		Generation using calibrator model Fluke 9500/9510		
c. Amplitude	Range	DC Signal ( $\pm$ )		Square Signal (10Hz to 10kHz)		Generation using calibrator model Fluke 9500/9510
		Into 50 $\Omega$	Into 1M $\Omega$	Into 50 $\Omega$	Into 1M $\Omega$	
	35 $\mu$ V to 1 mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 25 $\mu$ V	10mV/V + 1.1 $\mu$ V	10mV/V + 1.1 $\mu$ V	
	1mV to 21mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 25 $\mu$ V	2.4mV/V + 3.6 $\mu$ V	2.5mV/V + 1 $\mu$ V	
	21mV to 100mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 25 $\mu$ V	0.99mV/V + 1.7 $\mu$ V	0.99mV/V + 1.8 $\mu$ V	
	100mV to 556mV	0.25mV/V + 26 $\mu$ V	0.25mV/V + 45 $\mu$ V	0.98mV/V + 16 $\mu$ V	0.98mV/V + 16 $\mu$ V	
	556mV to 5.5V	0.25mV/V + 26 $\mu$ V	0.25mV/V + 45 $\mu$ V	0.44mV/V + 0.33mV	0.47mV/V + 0.3mV	
	5.5V to 10V	-	0.25mV/V + 45 $\mu$ V	-	0.47mV/V + 0.3mV	
	10V to 100V	-	0.25mV/V + 81 $\mu$ V	-	0.40mV/V + 35mV	
100V to 210V	-	0.25mV/V + 93 $\mu$ V	-	0.40mV/V + 35mV		
210V to 222V	-	0.25mV/V + 93 $\mu$ V	-	-		
<b>3. Source</b>						
DC VOLTAGE	0 to 200 mV 0.2 V to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	9.1 $\mu$ V/V + 0.11 $\mu$ V 7.6 $\mu$ V/V + 0.41 $\mu$ V 7.2 $\mu$ V/V + 2.1 $\mu$ V 12 $\mu$ V/V + 43 $\mu$ V 12 $\mu$ V/V + 42 $\mu$ V		Measurement using Multimeter model Wavetek 1281		
AC VOLTAGE	0 to 1000 V	(See Matrix G)		Measurement using Multimeter model Wavetek 1281		



## NO. SAMM 122

(Issue 3, 26 October 2010 replacement of  
SAMM 122 dated 5 October 2009)

Page 12 of 15

## FIELD OF CALIBRATION

## ELECTRICAL

## SITE CALIBRATION – CATEGORY 1

## SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks
<b>3. Source</b>			
DC CURRENT	0 to 200 $\mu$ A 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 0.2 A to 2 A 2 A to 10 A	0.12 mA/A + 0.45 nA 0.12 mA/A + 4.5 nA 0.12 mA/A + 45 nA 0.11 mA/A + 1.1 $\mu$ A 0.23 mA/A + 22 $\mu$ A 0.10 mA/A + 11 $\mu$ A	Measurement using Multimeter model Wavetek 1281
AC CURRENT	0 to 2 A	(See Matrix H)	Measurement using Multimeter model Wavetek 1281
	2 A to 10 A	(See Matrix H)	Measurement using Multimeter & Current Shunt model Wavetek 1281&4953
DC RESISTANCE	0 to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 0.2 k $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 0.2 M $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$ 0.2 G $\Omega$ to 1G $\Omega$	18 $\mu\Omega/\Omega$ + 22 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 65 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 0.63 m $\Omega$ 11 $\mu\Omega/\Omega$ + 6.3 m $\Omega$ 13 $\mu\Omega/\Omega$ + 56 m $\Omega$ 20 $\mu\Omega/\Omega$ + 1.3 $\Omega$ 37 $\mu\Omega/\Omega$ + 87 $\Omega$ 0.35 m $\Omega/\Omega$ + 10 k $\Omega$ 3.5 m $\Omega/\Omega$ + 0.5 M $\Omega$	Measurement using Multimeter model Wavetek 1281
FREQUENCY	100 mHz to 2.7 GHz	85 nHz/Hz	Measurement using Frequency Counter model Fluke PM6680B

## Signatory:

NO. NAME

I/C NO.

1. Lam Sik Lee

731122-07-5254



**NO. SAMM 122**(Issue 3, 26 October 2010 replacement of  
SAMM 122 dated 5 October 2009)

Page 13 of 15

**FIELD OF CALIBRATION****ELECTRICAL****SITE CALIBRATION – CATEGORY 1****SCOPE OF ACCREDITATION:****Matrix E - AC Voltage Measurement**

Range	Frequency						
	(10 to 31) Hz	(32 to 330) Hz	(0.3 to 10) kHz	(10 to 33) kHz	(30 to 100) kHz	(100 to 330) kHz	(0.300 to 1) MHz
0.09 mV to 2 mV	0.21 + 0.0051	0.20 + 0.0051	0.18 + 0.0051	0.27 + 0.0051	0.52 + 0.0051	1.1 + 0.021	2.2 + 0.024
0.2 mV to 20 mV	0.18 + 0.005	0.15 + 0.005	0.15 + 0.005	0.23 + 0.005	0.55 + 0.0049	1.1 + 0.021	2.3 + 0.024
20 mV to 200 mV	0.17 + 0.0077	0.13 + 0.0077	0.11 + 0.0074	0.21 + 0.0071	0.44 + 0.007	1.2 + 0.033	2.3 + 0.12
0.2 V to 2 V	0.12 + 0.026	0.073 + 0.017	0.074 + 0.0083	0.074 + 0.0082	0.13 + 0.017	0.39 + 0.084	2.1 + 0.34
2 V to 20 V	0.12 + 0.25	0.073 + 0.17	0.074 + 0.083	0.074 + 0.082	0.14 + 0.17	0.36 + 0.84	2.3 + 4.2
20 V to 200 V	0.17 + 2.6	0.11 + 1.7	0.074 + 0.82	0.083 + 1.7	0.27 + 2.5	0.77 + 42	8.2 + 110
200 V to 1000 V	0.2 + 17	0.2 + 17	0.16 + 17	0.19 + 33	0.28 + 10	-	-

The expanded uncertainties given in this table are expressed in **mV/V + mV****Matrix F - AC Current Measurement**

Range	Frequency			
	(0.01 to 3) kHz	(3 to 10) kHz	(10 to 20) kHz	(20 to 30) kHz
0 $\mu$ A to 32 $\mu$ A	0.79 + 1	1.1 + 2	2.3 + 6.8	2.8 + 10
32 $\mu$ A to 320 $\mu$ A	0.79 + 0.35	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
0.32 mA to 3.2 mA	0.79 + 0.48	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
3.2 mA to 32 mA	0.79 + 4.7	1.1 + 7.3	2.3 + 15	2.8 + 25
32 mA to 320 mA	0.91 + 36	1.1 + 55	2.3 + 73	2.8 + 110
0.32 A to 3.2 A	1.1 + 570	2.8 + 2900	-	-
3.2 A to 10.5 A	2.3 + 3400	5.7 + 11000	-	-

The expanded uncertainties given in this table are expressed in **mA/A +  $\mu$ A**

**NO. SAMM 122**(Issue 3, 26 October 2010 replacement of  
SAMM 122 dated 5 October 2009)

Page 14 of 15

**FIELD OF CALIBRATION****ELECTRICAL****SITE CALIBRATION – CATEGORY 1****SCOPE OF ACCREDITATION:****Matrix G - AC Voltage Source**

Range	Frequency							
	(10 – 40) Hz	(40 – 100) Hz	(0.1 – 2) kHz	(2 – 10) kHz	(10 – 30) kHz	(30 – 100) kHz	(100 – 300) kHz	(0.3 – 1) MHz
0 to 200mV	0.27+0.0044	0.25 +0.0045	0.23 +0.0023	0.23 +0.0045	0.48 +0.0089	0.90 +0.022	-	-
0.2V to 2V	0.20 +0.022	0.17 +0.022	0.15 +0.022	0.17 +0.022	0.28 +0.045	0.57 +0.23	3.4 +2.3	11 +23
2V to 20V	0.20 +0.22	0.17 +0.22	0.15 +0.22	0.17 +0.22	0.28 +0.45	0.57 +2.3	3.4 +2.3	11 +230
20V to 200V	0.20 +2.2	0.17 +2.2	0.15 +2.2	0.17 +2.2	0.28 +4.5	0.57 +23	3.4 +230	11 +2300
200V to 1000V	0.22 +10	0.20 +10	0.20 +10	0.20 +10	0.30 +22	0.60 +110	-	-

The expanded uncertainties given in this table are expressed in **mV/V + mV****Matrix H - AC Current Source**

Range	Frequency				
	(10 – 40) Hz	(40 – 300) Hz	(0.3 – 1) kHz	(1 – 5) kHz	(5 – 10) kHz
0 to 200 $\mu$ A	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	-
0.2 mA to 2 mA	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	-
2 mA to 20 mA	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	-
20 mA to 200 mA	0.36 + 22	0.36 + 22	0.36 + 22	0.36 + 22	-
0.2 A to 2 A	0.69 + 450	0.69 + 450	0.69 + 450	2.3 + 910	-
2 A to 10 A	0.63 + 110	0.61 + 100	0.61 + 48	0.61 + 100	0.61 + 100

The expanded uncertainties given in this table are expressed in **mA/A +  $\mu$ A**

# Schedule

Issue date : 26 October 2010  
Valid until : 21 October 2011



MS ISO/IEC 17025

**NO. SMM 122**

(Issue 3, 26 October 2010 replacement of SMM 122 dated 5 October 2009)

**FIELD OF CALIBRATION**

**TEMPERATURE**

**SCOPE OF ACCREDITATION:**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability expressed as an uncertainty ( $\pm$ )*	Remarks
<b>1. Temperature Indicator / Recorder / Controller</b>			
Type K	-250°C to -200°C -200°C to -100°C -100°C to +100°C +100°C to +600°C +600°C to +1372°C	0.66 °C 0.32 °C 0.23 °C 0.28 °C 0.32 °C	By electrical simulation using Wavetek 9100 Calibrator and reference table to ITS-90
Type J	-210°C to -100°C -100°C to +800°C +800°C to +1000°C +1000°C to +1200°C	0.30 °C 0.23 °C 0.26 °C 0.28 °C	By electrical simulation using Wavetek 9100 Calibrator and reference table to ITS-90
Type T	-250°C to -200°C -200°C to -100°C -100°C to 0°C 0°C to +400°C	0.68 °C 0.32 °C 0.27 °C 0.21 °C	By electrical simulation using Wavetek 9100 Calibrator and reference table to ITS-90
<b>2. Temperature Sensors</b>			
Thermocouple / Temperature probe	50°C to 250°C 250°C to 400°C 400°C to 650°C	0.30 °C 0.98 °C 3.5 °C	Comparison with Standard PT100 probe in calibration dry block
<b>3. Temperature &amp; Humidity Indicator</b>			
Thermohygraph / Thermohygraphometer	20 °C to 40°C 41°C to 60°C	0.68 °C 1.7 °C	Comparison with Vaisala HMM211 in Temperature Chamber
	30 %RH to 95 %RH	2.4 %RH	Comparison with Vaisala HMM211 in Humidity Chamber

**Signatory:**

**NO. NAME**

**I/C NO.**

1. Lam Sik Lee

731122-07-5254

