

NO: SAMM 122

LABORATORY LOCATION:
(PERMANENT LABORATORY)INTEGRAJAYA CALIBRATION TECHNOLOGIES
SDN. BHD.NO. 11, LINTANG BAYAN LEPAS 6,
TAMAN PERINDUSTRIAN BAYAN LEPAS,
FASA 4, 11900 BAYAN LEPAS,
PULAU PINANG, MALAYSIA

FIELD(S) OF CALIBRATION: ELECTRICAL, TEMPERATURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2005 (ISO/IEC 17025:2005).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

* The expanded uncertainties are based on an estimated confidence probability of approximately 95% and have a coverage factor of $k=2$ unless stated otherwise.

* The CMC is expressed as \pm (of indication in $\mu\text{V/V}$ + floor value in μV)

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
1. Indicating instrument			
DC VOLTAGE	0 μV to 200 mV 0.22 V to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	9.1 $\mu\text{V/V}$ + 0.68 μV 7.9 $\mu\text{V/V}$ + 1.1 μV 7.9 $\mu\text{V/V}$ + 4.0 μV 7.9 $\mu\text{V/V}$ + 7.4 μV 9.1 $\mu\text{V/V}$ + 91 μV 10 $\mu\text{V/V}$ + 0.57 mV	Generation using calibrator model Fluke 5700A
	1000 V to 10,000V	11 mV/V + 3.9 V	Compare with Vitrek 4670B Precision High Voltage Meter

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 2 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
1. Indicating instrument			
DC CURRENT	0 to 220 μ A 0.22 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 0.22 A to 2.2 A	57 μ A/A + 9.1 nA 51 μ A/A + 9.1 nA 57 μ A/A + 91 nA 68 μ A/A + 0.91 μ A 0.09 mA/A + 28 μ A	Generation using calibrator model Fluke 5700A
	0.32 A to 3.2 A 3.2 A to 10.5 A	0.60 mA/A + 0.44 mA 0.60 mA/A + 1.2 mA	Generation using calibrator model Wavetek 9100
	3.2 A to 32 A 32 A to 105 A 105 A to 200 A	1.3 mA/A + 50 mA 2.2 mA/A + 25 mA 2.3 mA/A + 29 mA	Generation using calibrator model Wavetek 9100 & 10 turn coil
	16 A to 160 A 160 A to 525 A 525 A to 1000 A	2.2 mA/A + 29 mA 2.3 mA/A + 23 mA 2.3 mA/A + 0.1 A	Generation using calibrator model Wavetek 9100 & 50 turn coil
AC VOLTAGE	0.22 mV to 220 V 220 V to 1050 V	(See Matrix A1) (See Matrix A2)	Generation using calibrator model Fluke 5700A & Wavetek 9100

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 3 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*					Remarks
AC CURRENT	0 μ A to 10.5 A	(See Matrix B)					Generation using calibrator model Wavetek 9100
	Range	FREQUENCY					Generation using calibrator model Fluke 5700A
		(10 to 20) Hz	(20 to 40) Hz	(0.04 to 1) kHz	(1 to 5) kHz	(5 to 10) kHz	
	2.2 μ A to 220 μ A	0.79 + 0.028	0.40 + 0.023	0.16 + 0.018	0.68 + 0.045	1.8 + 0.091	
	0.22 mA to 2.2 mA	0.79 + 0.045	0.40 + 0.04	0.16 + 0.04	0.68 + 0.45	1.8 + 0.91	
	2.2 mA to 22 mA	0.79 + 0.45	0.40 + 0.40	0.16 + 0.40	0.68 + 4.5	1.8 + 9.1	
	22 mA to 220 mA	0.79 + 4.5	0.40 + 4.0	0.16 + 4.0	0.68 + 45	1.8 + 91	
	0.22 A to 2.2 A	0.74 + 40	0.74 + 40	0.74 + 40	0.85 + 91	9.6 + 180	
	The expanded uncertainties given in this table are expressed in mA/A + μ A						
	Range	FREQUENCY					Generation using calibrator model Wavetek 9100 & 10/50 turn coil
(10 to 100) Hz		(100 to 440) Hz					
3.2 A to 32 A		2.2 mA/A + 48 mA	8.5 mA/A + 52 mA				
32 A to 200 A		3.2 mA/A + 96 mA	7.9 mA/A + 280 mA				
16 A to 160 A		3.1 mA/A + 43 mA					
160 A to 1000 A	3.2 mA/A + 440 mA						
**DC POWER	0.1024 mV to 10.5 kW	(See Matrix C)					Generation using calibrator model Wavetek 9100
**AC POWER	0.1024 mW to 7.7875 kW @ 10 Hz to 3 kHz	(See Matrix D) Power Factor = 1					Generation using calibrator model Wavetek 9100

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
1. Indicating instrument			
DC RESISTANCE (Specific Value)	1 m Ω 10 m Ω 100 m Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω	0.87 $\mu\Omega$ 11 $\mu\Omega$ 0.11 m Ω 1.1 m Ω 3.9 m Ω 32 m Ω 0.32 Ω 3.2 Ω 32 Ω	Generation using Standard Resistor model Agilent 42030A (4-Terminal Resistance)
	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	0.11 m Ω 0.20 m Ω 0.32 m Ω 0.58 m Ω 1.9 m Ω 3.7 m Ω 15 m Ω 28 m Ω 0.14 Ω 0.26 Ω 1.6 Ω 3.0 Ω 23 Ω 45 Ω 0.45 k Ω 1 k Ω 12 k Ω	Generation using calibrator model Fluke 5700A
DC RESISTANCE	0 Ω to 40 Ω 40 Ω to 400 Ω 400 Ω to 4 k Ω 4 k Ω to 40 k Ω 40 k Ω to 400 k Ω 400 k Ω to 4 M Ω 4 M Ω to 40 M Ω 40 M Ω to 400 M Ω	0.56 m Ω/Ω + 23 m Ω 0.17 m Ω/Ω + 25 m Ω 0.17 m Ω/Ω + 0.1 Ω 0.17 m Ω/Ω + 1 Ω 0.20 m Ω/Ω + 12 Ω 0.22 m Ω/Ω + 0.14 k Ω 0.56 m Ω/Ω + 2.4 k Ω 0.64 m Ω/Ω + 64 k Ω	Generation using calibrator model Wavetek 9100

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range		Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*						Remarks	
1. Indicating instrument										
DC RESISTANCE AT HIGH VOLTAGE RANGE (25V to 5KV)	Decade X1	1k Ω	10k Ω	100k Ω	1M Ω	10M Ω	100M Ω	1G Ω	Generation using Decade Resistance model IET HRSS	
	1	0.23 Ω	2.3 Ω	23 Ω	0.8 k Ω	9.8 k Ω	0.17 M Ω	4.1 M Ω		
	2	0.46 Ω	4.6 Ω	46 Ω	1.6 k Ω	59 k Ω	0.33 M Ω	9.9 M Ω		
	3	0.69 Ω	6.9 Ω	69 Ω	6.2 k Ω	62 k Ω	0.48 M Ω	14 M Ω		
	4	0.91 Ω	9.1 Ω	91 Ω	6.5 k Ω	65 k Ω	0.64 M Ω	17 M Ω		
	5	1.2 Ω	12 Ω	0.12 k Ω	7.0 k Ω	70 k Ω	0.80 M Ω	21 M Ω		
	6	1.4 Ω	14 Ω	0.14 k Ω	7.4 k Ω	74 k Ω	1.0 M Ω	25 M Ω		
	7	1.6 Ω	16 Ω	0.16 k Ω	8.0 k Ω	80 k Ω	1.2 M Ω	29 M Ω		
	8	1.9 Ω	19 Ω	0.19 k Ω	8.5 k Ω	86 k Ω	1.3 M Ω	33 M Ω		
	9	2.1 Ω	21 Ω	0.21 k Ω	9.2 k Ω	92 k Ω	1.5 M Ω	37 M Ω		
10	2.3 Ω	23 Ω	0.24 k Ω	9.8 k Ω	98 k Ω	1.7 M Ω	41 M Ω			
AC RESISTANCE (Specific Value)	Range	100 kHz	1 MHz	2 MHz	3 MHz	4 MHz	5 MHz	10 MHz	13 MHz	Generation using Standard Resistor model Agilent 42030A (4-Terminal Resistance Set)
	10 Ω	-	3.8 m Ω	5.6 m Ω	6.5 m Ω	7.5 m Ω	11 m Ω	41 m Ω	61 m Ω	
	100 Ω	-	39 m Ω	47 m Ω	56 m Ω	56 m Ω	56 m Ω	0.21 Ω	0.31 Ω	
	1 k Ω	0.39 Ω	0.39 Ω	0.39 Ω	0.39 Ω	0.47 Ω	0.56 Ω	2.1 Ω	3.1 Ω	
	10 k Ω	3.2 Ω	3.9 Ω	-	-	-	-	-	-	
100 k Ω	39 Ω	39 Ω	-	-	-	-	-	-		
CAPACITANCE	Range	Frequency						Generation using calibrator model Wavetek 9100		
		≤ 350 Hz			0.35 to 1.5 kHz					
	0.5 nF to 4.0 nF	3.4 mF/F + 17 pF			6.8 mF/F + 34 pF					
	4 nF to 40 nF	3.4 mF/F + 34 pF			6.8 mF/F + 68 pF					
	40 nF to 400 nF	3.4 mF/F + 0.18 nF			6.8 mF/F + 0.36 nF					
	400 nF to 4 μ F	4.5 mF/F + 1.8 nF			9.1 mF/F + 3.6 nF					
	4 μ F to 40 μ F	5.7 mF/F + 18 nF			11 mF/F + 36 nF					
	40 μ F to 400 μ F	5.7 mF/F + 0.18 μ F			11 mF/F + 0.36 μ F					
400 μ F to 4 mF	5.7 mF/F + 1.8 μ F			11 mF/F + 3.6 μ F						
4 mF to 40 mF	11 mF/F + 68 μ F			23 mF/F + 0.14 mF						

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range		Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*						Remarks
1. Indicating instrument									
CAPACITANCE (Specific Value)	Range	1 pF	10 pF	100 pF	1000 pF	10 nF	100 nF	1 uF	Generation using Standard Capacitor model Agilent 16380A & 16380C (4-Terminal Capacitor Set)
	120 Hz	-	-	-	-	1.2 pF	12 pF	0.13 nF	
	1 kHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	1.2 pF	12 pF	0.12 nF	
	10 kHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	1.2 pF	12 pF	0.12 nF	
	100 kHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	1.2 pF	12 pF	0.12 nF	
	1 MHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	-	-	-	
	2 MHz	0.72 fF	6.9 fF	0.069 pF	0.70 pF	-	-	-	
	3 MHz	0.80 fF	6.9 fF	0.069 pF	0.74 pF	-	-	-	
	4 MHz	0.93 fF	6.9 fF	0.069 pF	0.81 pF	-	-	-	
	5 MHz	1.2 fF	6.9 fF	0.070 pF	0.92 pF	-	-	-	
	10 MHz	2.6 fF	6.9 fF	0.076 pF	2.1 pF	-	-	-	
13 MHz	3.8 fF	7.0 fF	0.084 pF	2.9 pF	-	-	-		
INDUCTANCE	1 kHz 1mH to 900 mH			23 mH/H + 0.57 μ H				Generation using calibrator model IET LS- 400A	
INDUCTANCE (Specific Value)	100 Hz, 120 Hz, 1 kHz, 10 kHz, 100 kHz 100 μ H 1 mH 10 mH 100 mH 1 H			0.57 μ H 2.3 μ H 23 μ H 0.23 μ H 2.3 mH				Generation using calibrator model General Radio 1482	

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
FREQUENCY	0.1 Hz to 1 Hz 1 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz	9.8 nHz/Hz + 0.12 μ Hz 62 nHz/Hz + 2.2 μ Hz 81 nHz/Hz + 4.5 μ Hz 83 nHz/Hz + 26 μ Hz 83 nHz/Hz + 0.26 mHz 81 nHz/Hz + 4.4 mHz 81 nHz/Hz + 45 mHz 83 nHz/Hz + 0.23 Hz 85 nHz/Hz + 0.03 Hz	Using 10 MHz Reference model Fluke PM6680B
	0.1 Hz to 100 Hz 0.1 kHz to 100 kHz 0.1 MHz to 100 MHz 0.1 GHz to 9 GHz	1.2 μ Hz/Hz + 0.22 μ Hz 1.2 μ Hz/Hz + 4.1 nHz 34 nHz/Hz + 57 μ Hz 34 nHz/Hz + 7.2 μ Hz	Generation using Signal Generator model Agilent 33250A, E8663B & SRS FS725 Rubidium Standard (Reference)
2. Oscilloscope			
a) BANDWIDTH	100 mHz to 100 MHz 100.01 MHz to 550 MHz 550.01 MHz to 1.1 GHz	22 mHz/Hz 34 mHz/Hz 43 mHz/Hz	Generation using calibrator model Fluke 9500/9510
	1 GHz to 3 GHz 3 GHz to 6 GHz	34 mHz/Hz + 1.2 MHz 48 mHz/Hz + 0.35 MHz	Generation using Calibrator model Fluke 9500B & 9560
b) *TIME BASE	@100 mV to 500 mV 180.19 ps – 450 ps 450 ps – 900 ps 0.9 ns – 9 ns @100 mV to 1 V 9 ns - 100 ns 0.1 μ s - 10 μ s 10 μ s - 100 μ s 0.1 ms - 10 ms 10 ms - 100 ms 0.1 s - 10 s 10 s - 55 s	0.38 ns/s + 0.072 ps 0.77 ns/s + 0.079 ps 0.48 ns/s + 1.1 ps 0.063 ns/s + 79 ps 0.060 ns/s + 7.2 ns 0.063 ns/s + 79 ns 0.057 ns/s + 7.9 μ s 0.063 ns/s + 79 μ s 0.057 ns/s + 7.9 ms 0.046 ns/s + 57 ms	Generation using calibrator model Fluke 9500B & 9510/9560

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*				Remarks
2. Oscilloscope						
c) AMPLITUDE	Range	DC Signal (\pm)		Square Signal (10 Hz to 10 kHz)		Generation using calibrator model Fluke 9500 & 9510/9560
		Into 50 Ω	Into 1 M Ω	Into 50 Ω	Into 1 M Ω	
	35 μ V to 1 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 25 μ V	10 mV/V + 1.1 μ V	10 mV/V + 1.1 μ V	
	1 mV to 21 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 25 μ V	2.4 mV/V + 3.6 μ V	2.5 mV/V + 1 μ V	
	21 mV to 100 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 25 μ V	0.99 mV/V + 1.7 μ V	0.99 mV/V + 1.8 μ V	
	10.0 mV to 556 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 45 μ V	0.98 mV/V + 16 μ V	0.98 mV/V + 16 μ V	
	556 mV to 5.5 V	0.25 mV/V + 26 μ V	0.25 mV/V + 45 μ V	0.44 mV/V + 0.33 mV	0.47 mV/V + 0.3 mV	
	5.5 V to 10 V	-	0.25 mV/V + 45 μ V	-	0.47 mV/V + 0.3 mV	
	10 V to 100 V	-	0.25 mV/V + 81 μ V	-	0.40 mV/V + 35 mV	
	100 V to 210 V	-	0.25 mV/V + 93 μ V	-	0.40 mV/V + 35 mV	
210 V to 222 V	-	0.25 mV/V + 93 μ V	-	-		
3. Source						
DC VOLTAGE	0 mV to 100 mV 0.1 V to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	8.8 μ V/V + 0.49 μ V 8.2 μ V / V + 0.44 μ V 8.3 μ V/V + 0.93 μ V 11 μ V/V + 77 μ V 24 μ V/V + 29 μ V		Measurement using Multimeter model Agilent 3458A		
	1 kV to 10 kV	0.57 mV/V + 0.57 V		Measurement using High Voltage Meter model Vallaha 4600		
AC VOLTAGE	0 to 1000 V	(See Matrix E)		Measurement using Multimeter model Wavetek 1281		
	1 kV to 2 kV @ 20 Hz to 60 Hz	1.1 mV/V + 2.3 V		Measurement using High Voltage Meter model Vallaha 4600		
	2 kV to 15 kV @ 20 Hz to 60 Hz	5.7 mV/V + 57 V		Measurement using High Voltage Meter model Vallaha 4600		

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks			
3. Source						
DC CURRENT	0 nA to 100 nA 0.1 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 0.1 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 0.1 A to 1 A	40 μ A/A + 0.077 nA 29 μ A/A + 0.089 nA 29 μ A/A + 3.3 nA 29 μ A/A + 2.7 nA 29 μ A/A + 23 nA 29 μ A/A + 0.23 μ A 46 μ A/A + 2.3 μ A 0.13 mA/A + 26 μ A	Measurement using Multimeter model Agilent 3458A			
	1 A to 10 A 10 A to 20 A 20 A to 30 A	16 μ A/A + 0.40 mA 48 μ A/A + 0.25 mA 0.57 mA/A + 18 μ A	Measurement using Multimeter & Current shunt model Agilent 3458A & 34330A			
	30 A to 50 A 50 A to 100 A 100 A to 200 A 200 A to 300 A	46 μ A/A + 0.72 mA 60 μ A/A + 0.41 mA 71 μ A/A + 0.21 mA 71 μ A/A + 0.14 mA	Measurement using Multimeter & Current shunt model Agilent 3458A & Guildine 9230/1000			
AC CURRENT	0 to 2 A	(See Matrix F)	Measurement using Multimeter model Wavetek 1281			
	2 A to 10 A	(See Matrix F)	Measurement using Multimeter & Current Shunt model Wavetek 1281 & Wavetek 4953			
DC RESISTANCE	0 to 20 Ω 20 Ω to 200 Ω 0.2 k Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 0.2 M Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 0.2 G Ω to 1G Ω	18 $\mu\Omega/\Omega$ + 22 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 65 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 0.63 m Ω 11 $\mu\Omega/\Omega$ + 6.3 m Ω 13 $\mu\Omega/\Omega$ + 56 m Ω 20 $\mu\Omega/\Omega$ + 1.3 Ω 37 $\mu\Omega/\Omega$ + 87 Ω 0.35 m Ω/Ω + 10 k Ω 3.5 m Ω/Ω + 0.5 M Ω	Measurement using Multimeter model Wavetek 1281			
RF POWER	Range	FREQUENCY				Measurement using Power Meter & Power Sensor model Agilent EPM441A & 8482A
		(100 to 300) kHz	(0.3 to 1) MHz	(0.001 to 2) GHz	(2.0 to 4.2) GHz	
	-30 dBm to 20 dBm (1 μ W to 100 mW into 50 Ω)	4.6 %	3.5 %	3.3 %	3.7 %	

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*				Remarks
3. Source						
FREQUENCY	100 mHz to 1 Hz 1 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz 1 GHz to 2.7 GHz	83 nHz/Hz + 1.9 nHz 56 nHz/Hz + 2.9 μ Hz 85 nHz/Hz + 0.26 μ Hz 84 nHz/Hz + 11 μ Hz 83 nHz/Hz + 0.23 mHz 83 nHz/Hz + 1.9 mHz 84 nHz/Hz + 11 mHz 83 nHz/Hz + 0.23 Hz 83 nHz/Hz + 1.9 Hz 77 nHz/Hz + 31 Hz				Measurement using Frequency Counter model Fluke PM6680B
FREQUENCY	300 MHz – 12.4 GHz	17 nHz/Hz				Measurement using Frequency Counter & model Agilent 53132A and SRS FS725 Rubidium Standard (Reference)
**TIMER	1 s to 100 s 100 s to 400 s	38 us/s + 30 us 34 us/s + 1.8 ms				Measurement using DSO7104A Oscilloscope
**CAPACITANCE	Range	100 Hz	1 kHz	10 kHz	100 kHz	Measurement using HP 4263A LCR Meter
	1 pF to 10 pF	-	1.2 + 0.00012	5.0 + 0.000029	27 + .0000054	
	10 pF to 100 pF	1.9 + 0.00077	1.1 + 0.0013	2.0 + 0.00071	15 + 0.000099	
	100 pF to 1000 pF	1.80 + 0.0077	1.1 + 0.013	1.9 + 0.0073	14 + 0.0011	
	1 nF to 10 nF	1.8 + 0.077	1.1 + 0.13	1.9 + 0.074	14 + 0.011	
	10 nF to 100 nF	2.9 + 0.50	2.9 + 0.50	2.7 + 0.53	16 + 0.094	
	100 nF to 1000 nF	3.1 + 4.6	2.0 + 1.2	3.3 + 4.4	19 + 0.79	
	1 uF to 10 uF	5.8 + 25	5.4 + 27	5.4 + 27	31 + 4.7	
	10 uF to 100 uF	13 + 130	8.8 + 170	21 + 700	-	
	100 uF to 1000 uF	73 + 200	47 + 310	-	-	
	1 mF to 10 mF	-	55 + 490000	-	-	
	10 mF to 100 mF	-	55 + 490000	-	-	
The expanded uncertainties given in this table expressed in mF/F + pF						

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Signatories:

- Lam Sik Lee
- Mohd Rahimi Bin Samsuddin (* EOS: Time base only, ** Excluded for EOS)

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL**Matrix A1 - AC Voltage Measurement**

Range	FREQUENCY							
	(10 to 20) Hz	(20 to 40) Hz	(0.04 to 20) kHz	(20 to 50) kHz	(50 to 100) kHz	(100 to 300) kHz	(300 to 500) kHz	(0.5 to 1.0) MHz
0.22 mV to 2.2 mV	0.62 + 0.0051	0.24 + 0.0051	0.12 + 0.0051	0.42 + 0.0051	0.96 + 0.0079	1.2 + 0.015	1.9 + 0.028	3.9 + 0.028
2.2 mV to 22 mV	0.62 + 0.0057	0.24 + 0.0057	0.12 + 0.0057	0.42 + 0.0057	0.96 + 0.0079	1.2 + 0.014	1.9 + 0.028	3.9 + 0.028
22 mV to 220 mV	0.62 + 0.015	0.24 + 0.0091	0.12 + 0.0091	0.36 + 0.0091	0.96 + 0.028	1.2 + 0.028	1.9 + 0.040	3.9 + 0.12
0.22 V to 2.2 V	0.57 + 0.091	0.18 + 0.028	0.085 + 0.0068	0.14 + 0.018	0.28 + 0.079	0.49 + 0.15	1.2 + 0.040	2.5 + 0.96
2.2 V to 22 V	0.57 + 0.91	0.18 + 0.28	0.085 + 0.068	0.14 + 0.18	0.28 + 0.040	0.57 + 1.7	1.4 + 4.9	3.1 + 9.6
22 V to 220 V	0.57 + 9.1	0.18 + 2.8	0.091 + 0.91	0.25 + 4.0	0.57 + 9.1	1.7 + 100	5.3 + 100	13 + 220

The expanded uncertainties given in this table are expressed in mV/V + mV

Matrix A2 - AC Voltage Measurement

Range	FREQUENCY					
	(15 to 50) Hz	(0.05 to 1) kHz	(1 to 3) kHz	(3 to 10) kHz	(10 to 20) kHz	(20 to 30) kHz
220 V to 1100 V	0.45 + 18	0.091 + 4.0	-	-	-	-
220 V to 320 V	-	-	0.91 + 22	0.91 + 36	1.4 + 48	1.7 + 73
320 V to 800 V	-	-	0.91 + 71	0.91 + 120	1.4 + 180	1.7 + 240
800 V to 1050 V	-	-	0.90 + 150	0.91 + 240	1.4 + 360	-

Matrix B - AC Current Measurement

Range	Frequency			
	(0.01 to 3) kHz	(3 to 10) kHz	(10 to 20) kHz	(20 to 30) kHz
0 µA to 32 µA	0.79 + 1	1.1 + 2	2.3 + 6.8	2.8 + 10
32 µA to 320 µA	0.79 + 0.35	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
0.32 mA to 3.2 mA	0.79 + 0.48	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
3.2 mA to 32 mA	0.79 + 4.7	1.1 + 7.3	2.3 + 15	2.8 + 25
32 mA to 320 mA	0.91 + 36	1.1 + 55	2.3 + 73	2.8 + 110
0.32 A to 3.2 A	1.1 + 570	2.8 + 2900	-	-
3.2 A to 10.5 A	2.3 + 3400	5.7 + 11000	-	-

The expanded uncertainties given in this table are expressed in mA/A + µA

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 12 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Matrix C - DC Power Measurement

Voltage	Current	DC Power	CMC
320 mV	3.2 mA	(0.1024 to 1.024) mW	0.041 + 0.56
	32 mA	(1.024 to 10.24) mW	0.17 + 0.42
	320 mA	(10.24 to 102.4) mW	0.20 + 4.1
	3.2 A	(0.1024 to 1.024) W	0.70 + 26
	10.5 A	(1.024 to 3.36) W	0.62 + 430
3.2 V	3.2 mA	(1.024 to 10.24) mW	0.17 + 0.42
	32 mA	(10.24 to 102.4) mW	0.17 + 4.2
	320 mA	(0.1024 to 1.024) W	0.20 + 41
	3.2 A	(1.024 to 10.24) W	0.7 + 260
	10.5 A	(10.24 to 33.6) W	0.62 + 4200
32 V	3.2 mA	(10.24 to 102.4) mW	0.17 + 4.2
	32 mA	(0.1024 to 1.024) W	0.17 + 42
	320 mA	(1.024 to 10.24) W	0.2 + 410
	3.2 A	(10.24 to 102.4) W	0.7 + 2600
	10.5 A	(102.4 to 336.0) W	0.19 + 550000
320 V	3.2 mA	(0.1024 to 1.024) W	0.042 + 560
	32 mA	(1.024 to 10.24) W	0.16 + 640
	320 mA	(10.24 to 102.4) W	0.2 + 4100
	3.2 A	(0.1024 to 1.024) kW	0.7 + 26000
	10.5 A	(1.024 to 3.36) kW	0.62 + 430000
1000 V	3.2 mA	(0.32 to 3.2) W	0.11 + 520
	32 mA	(3.2 to 32) W	0.11 + 5200
	320 mA	(32 to 320) W	0.13 + 51000
	3.2 A	(0.32 to 3.2) kW	0.62 + 370000
	10.5 A	(3.2 to 10.5) kW	0.66 + 800000

The expanded uncertainties given in this table expressed in mW/W + uW

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 13 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

Matrix D - AC Power Measurement

Voltage	Current	AC Power	(10 to 40) Hz	(40 to 100) Hz	(0.1 to 1) kHz	(1 to 3) kHz
320 mV	3.2 mA	(0.1024 to 1.024) mW	0.97 + 77 nW	0.97 + 77 nW	0.97 + 77 nW	0.97 + 77 nW
	32 mA	(1.024 to 10.24) mW	1.0 + 0.00066	1.0 + 0.00066	1.0 + 0.00066	1.0 + 0.00066
	320 mA	(10.24 to 102.4) mW	1.1 + 0.0064	1.1 + 0.0064	1.1 + 0.0064	1.1 + 0.0064
	3.2 A	(0.1024 to 1.024) W	1.4 + 0.081	1.4 + 0.081	1.4 + 0.081	1.4 + 0.081
	10.5 A	(1.024 to 3.36) W	2.5 + 0.63	2.5 + 0.63	2.5 + 0.63	2.5 + 0.63
3.2 V	3.2 mA	(1.024 to 10.24) mW	0.99 + 0.00065	0.99 + 0.00065	0.99 + 0.00065	0.99 + 0.00065
	32 mA	(10.24 to 102.4) mW	0.99 + 0.0066	0.99 + 0.0066	0.99 + 0.0066	0.99 + 0.0066
	320 mA	(0.1024 to 1.024) W	1.1 + 0.075	1.1 + 0.075	1.1 + 0.075	1.1 + 0.075
	3.2 A	(1.024 to 10.24) W	1.4 + 0.71	1.4 + 0.71	1.4 + 0.71	1.4 + 0.71
	10.5 A	(10.24 to 33.6) W	2.5 + 6.3	2.5 + 6.3	2.5 + 6.3	2.5 + 6.3
32 V	3.2 mA	(10.24 to 102.4) mW	0.98 + 0.0068	0.98 + 0.0068	0.98 + 0.0068	0.98 + 0.0068
	32 mA	(0.1024 to 1.024) W	0.97 + 0.083	0.97 + 0.083	0.97 + 0.083	0.97 + 0.083
	320 mA	(1.024 to 10.24) W	1.1 + 0.67	1.1 + 0.67	1.1 + 0.67	1.1 + 0.67
	3.2 A	(10.24 to 102.4) W	1.4 + 7.4	1.4 + 7.4	1.4 + 7.4	1.4 + 7.4
	10.5 A	(102.4 to 336.0) W	2.5 + 63	2.5 + 63	2.5 + 63	2.5 + 63
105 V	3.2 mA	(0.0336 to 0.336) W	0.83 + 0.08	0.83 + 0.08	0.83 + 0.08	0.83 + 0.08
	32 mA	(0.336 to 3.36) W	0.84 + 0.8	0.84 + 0.8	0.84 + 0.8	0.84 + 0.8
	320 mA	(3.36 to 33.6) W	0.94 + 7.7	0.94 + 7.7	0.94 + 7.7	0.94 + 7.7
	3.2 A	(0.0336 to 0.336) kW	1.2 + 73	1.2 + 73	1.2 + 73	1.2 + 73
	10.5 A	(0.1103 to 1.103) kW	2.6 + 110	2.6 + 110	2.6 + 110	2.6 + 110
320 V	3.2 mA	(0.1024 to 1.024) W	-	1.0 + 0.068	1.1 + 0.11	1.3 + 0.059
	32 mA	(1.024 to 10.24) W	-	1.1 + 0.56	1.1 + 0.56	1.3 + 0.49
	320 mA	(10.24 to 102.4) W	-	1.2 + 5.5	1.2 + 5.5	1.4 + 4.8
	3.2 A	(0.1024 to 1.024) kW	-	1.4 + 73	1.4 + 73	1.6 + 67
	10.5 A	(1.024 to 3.36) kW	-	2.5 + 620	2.5 + 620	2.6 + 610
750 V	3.2 mA	(0.24 to 2.4) W	-	0.81 + 0.87	0.81 + 0.87	1.1 + 0.81
	32 mA	(2.4 to 24) W	-	0.81 + 8.7	0.81 + 8.7	1.1 + 8.2
	320 mA	(24 to 240) W	-	1.2 + 11	1.2 + 11	1.4 + 12
	3.2 A	(0.24 to 2.4) kW	-	1.2 + 720	1.2 + 720	1.4 + 670
	10.5 A	(0.7875 to 7.7875) kW	-	2.6 + 830	2.6 + 830	2.7 + 830

The expanded uncertainties given in this table expressed in mW/W + mW

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL**Matrix E - AC Voltage Source**

Range	FREQUENCY							
	(10 – 40) Hz	(40 – 100) Hz	(0.1 – 2) kHz	(2 – 10) kHz	(10 – 30) kHz	(30 – 100) kHz	(100 – 300) kHz	(0.3 – 1) MHz
0 to 200 mV	0.27 + 0.0044	0.25 + 0.0045	0.23 + 0.0023	0.23 + 0.0045	0.48 + 0.0089	0.90 + 0.022	-	-
0.2 V to 2 V	0.20 + 0.022	0.17 + 0.022	0.15 + 0.022	0.17 + 0.022	0.28 + 0.045	0.57 + 0.23	3.4 + 2.3	11 + 23
2 V to 20 V	0.20 + 0.22	0.17 + 0.22	0.15 + 0.22	0.17 + 0.22	0.28 + 0.45	0.57 + 2.3	3.4 + 2.3	11 + 230
20 V to 200 V	0.20 + 2.2	0.17 + 2.2	0.15 + 2.2	0.17 + 2.2	0.28 + 4.5	0.57 + 23	3.4 + 230	11 + 2300
200 V to 1000 V	0.22 + 10	0.20 + 10	0.20 + 10	0.20 + 10	0.30 + 22	0.60 + 110	-	-

The expanded uncertainties given in this table are expressed in mV/V + mV

Matrix F - AC Current Source

Range	FREQUENCY				
	(10 – 40) Hz	(40 – 300) Hz	(0.3 – 1) kHz	(1 – 5) kHz	(5 – 10) kHz
0 to 200 μ A	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	-
0.2 mA to 2 mA	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	-
2 mA to 20 mA	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	-
20 mA to 200 mA	0.36 + 22	0.36 + 22	0.36 + 22	0.36 + 22	-
0.2 A to 2 A	0.69 + 450	0.69 + 450	0.69 + 450	2.3 + 910	-
2 A to 10 A	0.63 + 110	0.61 + 100	0.61 + 48	0.61 + 100	0.61 + 100

The expanded uncertainties given in this table are expressed in mA/A + μ A

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 15 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Indicating instrument			
DC VOLTAGE	0 μ V to 220 mV 0.22 V to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	9.1 μ V/V + 0.68 μ V 7.9 μ V/V + 1.1 μ V 7.9 μ V/V + 4.0 μ V 7.9 μ V/V + 7.4 μ V 9.1 μ V/V + 91 μ V 10 μ V/V + 0.57 mV	Generation using calibrator model Fluke 5700A
	1000 V to 10,000 V	11 mV/V + 3.9 V	Compare with Vitrek 4670B Precision High Voltage Meter
DC CURRENT	0 to 220 μ A 0.22 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 0.22 A to 2.2 A	57 μ A/A + 9.1 nA 51 μ A/A + 9.1 nA 57 μ A/A + 91 nA 68 μ A/A + 0.91 μ A 0.09 mA/A + 28 μ A	Generation using calibrator model Fluke 5700A
	0.32 A to 3.2 A 3.2 A to 10.5 A	0.6 mA/A + 0.44 mA 0.6 mA/A + 1.2 mA	Generation using calibrator model Wavetek 9100
	0 A to 500 A	5 mA/A + 0.51 A	Generation using calibrator model Fluke 5500A & 50 turn coil
AC VOLTAGE	0.22 mV to 220 V 220 V to 1050 V	(See Matrix G1) (See Matrix G2)	Generation using calibrator model Fluke 5700A & Wavetek 9100

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 16 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*					Remarks
1. Indicating instrument							
AC CURRENT	0 μ A to 10.5 A	(See Matrix H)					Generation using calibrator model Wavetek 9100
		Frequency					Generation using calibrator model Fluke 5700A
	Range	(10 to 20) Hz	(20 to 40) Hz	(0.04 to 1) kHz	(1 to 5) kHz	(5 to 10) kHz	
	2.2 μ A to 220 μ A	0.79 + 0.028	0.40 + 0.023	0.16 + 0.018	0.68 + 0.045	1.8 + 0.091	
	0.22 mA to 2.2 mA	0.79 + 0.045	0.40 + 0.04	0.16 + 0.04	0.68 + 0.45	1.8 + 0.91	
	2.2 mA to 22 mA	0.79 + 0.45	0.40 + 0.40	0.16 + 0.40	0.68 + 4.5	1.8 + 9.1	
	22 mA to 220 mA	0.79 + 4.5	0.40 + 4.0	0.16 + 4.0	0.68 + 45	1.8 + 91	
	0.22 A to 2.2 A	0.74 + 40	0.74 + 40	0.74 + 40	0.85 + 91	9.6 + 180	
The expanded uncertainties given in this table are expressed in mA/A + μ A							
	110 A to 200 A @ 65 Hz to 440 Hz	5.1 mA/A + 0.51 A					Generation using calibrator model Fluke 5500A & 50 turn coil.
	110 A to 500 A @ 45 Hz to 65 Hz	5 mA/A + 0.5 A					

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 17 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
**DC POWER	0.0109 mW to 11 kW	(See Matrix I)	Generation using Calibrator model Fluke 5500A
**AC POWER	0.0109 mW to 11 kW @ 10 Hz to 10 kHz	(See Matrix J) Power Factor = 1	Generation using Calibrator model Fluke 5500A
DC RESISTANCE (Specific Value)	1 m Ω 10 m Ω 100 m Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω	0.87 $\mu\Omega$ 11 $\mu\Omega$ 0.11 m Ω 1.1 m Ω 3.9 m Ω 32 m Ω 0.32 Ω 3.2 Ω 32 Ω	Generation using Standard Resistor model Agilent 42030A (4-Terminal Resistance set)
	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	0.11 m Ω 0.20 m Ω 0.32 m Ω 0.58 m Ω 1.9 m Ω 3.7 m Ω 15 m Ω 28 m Ω 0.14 Ω 0.26 Ω 1.6 Ω 3.0 Ω 23 Ω 45 Ω 0.45 k Ω 1 k Ω 12 k Ω	Generation using calibrator model Fluke 5700A

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 18 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range		Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*								Remarks
DC RESISTANCE	0 Ω to 40 Ω 40 Ω to 400 Ω 400 Ω to 4 k Ω 4 k Ω to 40 k Ω 40 k Ω to 400 k Ω 400 k Ω to 4 M Ω 4 M Ω to 40 M Ω 40 M Ω to 400 M Ω		0.56 m Ω / Ω + 23 m Ω 0.17 m Ω / Ω + 25 m Ω 0.17 m Ω / Ω + 0.1 Ω 0.17 m Ω / Ω + 1 Ω 0.20 m Ω / Ω + 12 Ω 0.22 m Ω / Ω + 0.14 k Ω 0.56 m Ω / Ω + 2.4 k Ω 0.64 m Ω / Ω + 64 k Ω								Generation using calibrator model Wavetek 9100
DC RESISTANCE AT HIGH VOLTAGE (25 V to 5 kV)	Decade X1	1k Ω	10k Ω	100k Ω	1M Ω	10M Ω	100M Ω	1G Ω	Generation using Decade Resistance model IET HRRS		
	1	0.23 Ω	2.3 Ω	23 Ω	0.8 k Ω	9.8 k Ω	0.17 M Ω	4.1 M Ω			
	2	0.46 Ω	4.6 Ω	46 Ω	1.6 k Ω	59 k Ω	0.33 M Ω	9.9 M Ω			
	3	0.69 Ω	6.9 Ω	69 Ω	6.2 k Ω	62 k Ω	0.48 M Ω	14 M Ω			
	4	0.91 Ω	9.1 Ω	91 Ω	6.5 k Ω	65 k Ω	0.64 M Ω	17 M Ω			
	5	1.2 Ω	12 Ω	0.12 k Ω	7.0 k Ω	70 k Ω	0.80 M Ω	21 M Ω			
	6	1.4 Ω	14 Ω	0.14 k Ω	7.4 k Ω	74 k Ω	1.0 M Ω	25 M Ω			
	7	1.6 Ω	16 Ω	0.16 k Ω	8.0 k Ω	80 k Ω	1.2 M Ω	29 M Ω			
	8	1.9 Ω	19 Ω	0.19 k Ω	8.5 k Ω	86 k Ω	1.3 M Ω	33 M Ω			
	9	2.1 Ω	21 Ω	0.21 k Ω	9.2 k Ω	92 k Ω	1.5 M Ω	37 M Ω			
10	2.3 Ω	23 Ω	0.24 k Ω	9.8 k Ω	98 k Ω	1.7 M Ω	41 M Ω				
AC RESISTANCE (Specific Value)	Range	100 kHz	1 MHz	2 MHz	3 MHz	4 MHz	5 MHz	10 MHz	13 MHz	Generation using Standard Resistor model Agilent 42030A (4-Terminal Resistance set)	
	10 Ω	-	3.8 m Ω	5.6 m Ω	6.5 m Ω	7.5 m Ω	11 m Ω	41 m Ω	61 m Ω		
	100 Ω	-	39 m Ω	47 m Ω	56 m Ω	56 m Ω	56 m Ω	0.21 Ω	0.31 Ω		
	1 k Ω	0.39 Ω	0.39 Ω	0.39 Ω	0.39 Ω	0.47 Ω	0.56 Ω	2.1 Ω	3.1 Ω		
	10 k Ω	3.2 Ω	3.9 Ω	-	-	-	-	-	-		
100 k Ω	39 Ω	39 Ω	-	-	-	-	-	-			

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 19 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*		Remarks					
1. Indicating instrument									
CAPACITANCE	Range	Frequency		Generation using calibrator model Wavetek 9100					
		≤ 350 Hz	0.35 to 1.5 kHz						
	0.5 nF to 4.0 nF	3.4 mF/F + 17 pF	3.4 mF/F + 17 pF						
	4 nF to 40 nF	3.4 mF/F + 34 pF	3.4 mF/F + 34 pF						
	40 nF to 400 nF	3.4 mF/F + 0.18 nF	3.4 mF/F + 0.18 nF						
	400 nF to 4 μ F	4.5 mF/F + 1.8 nF	4.5 mF/F + 1.8 nF						
	4 μ F to 40 μ F	5.7 mF/F + 18 nF	5.7 mF/F + 18 nF						
	40 μ F to 400 μ F	5.7 mF/F + 0.18 μ F	5.7 mF/F + 0.18 μ F						
	400 μ F to 4 mF	5.7 mF/F + 1.8 μ F	5.7 mF/F + 1.8 μ F						
4 mF to 40 mF	11 mF/F + 68 μ F	11 mF/F + 68 μ F							
CAPACITANCE (Specific Value)	Range	1 pF	10 pF	100 pF	1000 pF	10 nF	100 nF	1 μ F	Generation using Standard Capacitor model Agilent 16380A & 16380C (4-Terminal Capacitor Set)
	120 Hz	-	-	-	-	1.2 pF	12 pF	0.13 nF	
	1 kHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	1.2 pF	12 pF	0.12 nF	
	10 kHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	1.2 pF	12 pF	0.12 nF	
	100 kHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	1.2 pF	12 pF	0.12 nF	
	1 MHz	0.69 fF	6.9 fF	0.069 pF	0.69 pF	-	-	-	
	2 MHz	0.72 fF	6.9 fF	0.069 pF	0.70 pF	-	-	-	
	3 MHz	0.80 fF	6.9 fF	0.069 pF	0.74 pF	-	-	-	
	4 MHz	0.93 fF	6.9 fF	0.069 pF	0.81 pF	-	-	-	
	5 MHz	1.2 fF	6.9 fF	0.070 pF	0.92 pF	-	-	-	
	10 MHz	2.6 fF	6.9 fF	0.076 pF	2.1 pF	-	-	-	
13 MHz	3.8 fF	7.0 fF	0.084 pF	2.9 pF	-	-	-		

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 20 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Indicating instrument			
FREQUENCY	0.1 Hz to 1 Hz 1 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz	9.8 nHz/Hz + 0.12 μ Hz 62 nHz/Hz + 2.2 μ Hz 81 nHz/Hz + 4.5 μ Hz 83 nHz/Hz + 26 μ Hz 83 nHz/Hz + 0.26 mHz 81 nHz/Hz + 4.4 mHz 81 nHz/Hz + 45 mHz 83 nHz/Hz + 0.23 Hz 85 nHz/Hz + 0.03 Hz	Using 10 MHz Reference model Fluke PM6680B
	0.1 Hz to 100 Hz 0.1 kHz to 100 kHz 0.1 MHz to 100 MHz 0.1 GHz to 9 GHz	1.2 μ Hz/Hz + 0.22 μ Hz 1.2 μ Hz/Hz + 4.1 nHz 34 nHz/Hz + 57 μ Hz 34 nHz/Hz + 7.2 μ Hz	Generation using Signal Generator model Agilent 33250A, E8663B & SRS FS725 Rubidium Standard (Reference)
2. Oscilloscope			
a. BANDWIDTH	100 mHz to 100 MHz 100.01 MHz to 550 MHz 550.01 MHz to 1.1 GHz	22 mHz/Hz 34 mHz/Hz 43 mHz/Hz	Generation using calibrator model Fluke 9500/9510
	1 GHz to 3 GHz 3 GHz to 6 GHz	34 mHz/Hz + 1.2 MHz 48 mHz/Hz + 0.35 MHz	Generation using Oscilloscope Calibrator model Fluke 9500B & 9560
b. *TIME BASE	@100 mV to 500 mV 180.19 ps – 450 ps 450 ps – 900 ps 0.9 ns – 9 ns @100 mV to 1 V 9 ns - 100 ns 0.1 μ s - 10 μ s 10 μ s - 100 μ s 0.1 ms - 10 ms 10 ms - 100 ms 0.1 s - 10 s 10 s - 55 s	0.38 ns/s + 0.072 ps 0.77 ns/s + 0.079 ps 0.48 ns/s + 1.1 ps 0.063 ns/s + 79 ps 0.060 ns/s + 7.2 ns 0.063 ns/s + 79 ns 0.057 ns/s + 7.9 μ s 0.063 ns/s + 79 μ s 0.057 ns/s + 7.9 ms 0.046 ns/s + 57 ms	Generation using calibrator model Fluke 9500/9560

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 21 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*		Remarks
c. AMPLITUDE	Range	DC Signal (\pm)		Generation using calibrator model Fluke 9500 & 9510/9560
		Into 50 Ω	Into 1 M Ω	
	35 μ V to 1 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 25 μ V	
	1 mV to 21 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 25 μ V	
	21 mV to 100 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 25 μ V	
	100 mV to 556 mV	0.25 mV/V + 26 μ V	0.25 mV/V + 45 μ V	
	556 mV to 5.5 V	0.25 mV/V + 26 μ V	0.25 mV/V + 45 μ V	
	5.5 V to 10 V	-	0.25 mV/V + 45 μ V	
	10 V to 100 V	-	0.25 mV/V + 81 μ V	
	100 V to 210 V	-	0.25 mV/V + 93 μ V	
	210 V to 222 V	-	0.25 mV/V + 93 μ V	
	Range	Square Signal (10 Hz to 10 kHz)		
		Into 50 Ω	Into 1 M Ω	
	35 μ V to 1 mV	10 mV/V + 1.1 μ V	10 mV/V + 1.1 μ V	
	1 mV to 21 mV	2.4 mV/V + 3.6 μ V	2.5 mV/V + 1 μ V	
	21 mV to 100 mV	0.99 mV/V + 1.7 μ V	0.99 mV/V + 1.8 μ V	
	100 mV to 556 mV	0.98 mV/V + 16 μ V	0.98 mV/V + 16 μ V	
	556 mV to 5.5 V	0.44 mV/V + 0.33 mV	0.47 mV/V + 0.3 mV	
	5.5 V to 10 V	-	0.47 mV/V + 0.3 mV	
	10 V to 100 V	-	0.40 mV/V + 35 mV	
100 V to 210 V	-	0.40 mV/V + 35 mV		
210 V to 222 V	-	-		

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 22 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
3. Source			
DC VOLTAGE	0 mV to 100 mV 0.1 V to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	8.8 μ V/V + 0.49 μ V 8.2 μ V/V + 0.44 μ V 8.3 μ V/V + 0.93 μ V 11 μ V/V + 77 μ V 24 μ V/V + 29 μ V	Measurement using Multimeter model Agilent 3458A
	1 kV to 10 kV	0.57 mV/V + 0.57 V	Measurement using High Voltage Meter model Valhalla 4600
AC VOLTAGE	0 to 1000 V	(See Matrix K)	Measurement using Multimeter model Wavetek 1281
	1 kV to 2 kV @ 20 Hz to 60 Hz 2 kV to 15 kV @ 20 Hz to 60 Hz	1.1 mV/V + 2.3 V 5.7 mV/V + 57 V	Measurement using High Voltage Meter model Valhalla 4600
DC CURRENT	0 nA to 100 nA 0.1 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 0.1 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 0.1 A to 1 A	40 μ A/A + 0.077 nA 29 μ A/A + 0.089 nA 29 μ A/A + 3.3 nA 29 μ A/A + 2.7 nA 29 μ A/A + 23 nA 29 μ A/A + 0.23 μ A 46 μ A/A + 2.3 μ A 0.13 mA/A + 26 μ A	Measurement using Multimeter model Agilent 3458A
	1 A to 10 A 10 A to 20 A 20 A to 30 A	16 μ A/A + 0.40 mA 48 μ A/A + 0.25 mA 0.57 mA/A + 18 μ A	Measurement using Multimeter & Current shunt model Agilent 3458A & 34330A
	30 A to 50 A 50 A to 100 A 100 A to 200 A 200 A to 300 A	46 μ A/A + 0.72 mA 60 μ A/A + 0.41 mA 71 μ A/A + 0.21 mA 71 μ A/A + 0.14 mA	Measurement using Multimeter & Current shunt model Agilent 3458A & Guildine9230/1000

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 23 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks			
3. Source						
AC CURRENT	0 to 2 A	(See Matrix L)	Measurement using Multimeter model Wavetek 1281			
	2 A to 10 A	(See Matrix L)	Measurement using Multimeter & Current Shunt model Wavetek 1281&4953			
DC RESISTANCE	0 to 20 Ω 20 Ω to 200 Ω 0.2 k Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 0.2 M Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 0.2 G Ω to 1G Ω	18 $\mu\Omega/\Omega$ + 22 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 65 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 0.63 m Ω 11 $\mu\Omega/\Omega$ + 6.3 m Ω 13 $\mu\Omega/\Omega$ + 56 m Ω 20 $\mu\Omega/\Omega$ + 1.3 Ω 37 $\mu\Omega/\Omega$ + 87 Ω 0.35 m Ω/Ω + 10 k Ω 3.5 m Ω/Ω + 0.5 M Ω	Measurement using Multimeter model Wavetek 1281			
RF POWER	Range -30 dBm to 20 dBm (1 μ W to 100 mW into 50 Ω)	FREQUENCY				Measurement using Power Meter & Power Sensor model Agilent EPM441A & 8482A
		(100 to 300) kHz	(0.3 to 1) MHz	(0.001 to 2) GHz	(2.0 to 4.2) GHz	
		4.6 %	3.5 %	3.3 %	3.7 %	
FREQUENCY	100 mHz to 1 Hz 1 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz 1 GHz to 2.7 GHz	83 nHz/Hz + 1.9 nHz 56 nHz/Hz + 2.9 μ Hz 85 nHz/Hz + 0.26 μ Hz 84 nHz/Hz + 11 μ Hz 83 nHz/Hz + 0.23 mHz 83 nHz/Hz + 1.9 mHz 84 nHz/Hz + 11 mHz 83 nHz/Hz + 0.23 Hz 83 nHz/Hz + 1.9 Hz 77 nHz/Hz + 31 Hz	Measurement using Frequency Counter model Fluke PM6680B			
	300 MHz – 12.4 GHz	17 nHz/Hz	Measurement using Frequency Counter & model Agilent 53132A and SRS FS725 Rubidium Standard (Reference)			

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 24 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (±)*					Remarks
**TIMER	1 s to 100 s 100 s to 400 s	38 us/s + 30 us 34 us/s + 1.8 ms					Measurement using DSO7104A Oscilloscope
**CAPACITANCE	Range	100 Hz	1 kHz	10 kHz	100 kHz	Measurement using HP 4263A LCR Meter	
	1 pF to 10 pF	-	1.2 + 0.00012	5.0 + 0.000029	27 + 0.0000054		
	10 pF to 100 pF	1.9 + 0.00077	1.1 + 0.0013	2.0 + 0.00071	15 + 0.000099		
	100 pF to 1000 pF	1.80 + 0.0077	1.1 + 0.013	1.9 + 0.0073	14 + 0.0011		
	1 nF to 10 nF	1.8 + 0.077	1.1 + 0.13	1.9 + 0.074	14 + 0.011		
	10 nF to 100 nF	2.9 + 0.50	2.9 + 0.50	2.7 + 0.53	16 + 0.094		
	100 nF to 1000 nF	3.1 + 4.6	2.0 + 1.2	3.3 + 4.4	19 + 0.79		
	1 uF to 10 uF	5.8 + 25	5.4 + 27	5.4 + 27	31 + 4.7		
	10 uF to 100 uF	13 + 130	8.8 + 170	21 + 700	-		
	100 uF to 1000 uF	73 + 200	47 + 310	-	-		
	1 mF to 10 mF	-	55 + 490000	-	-		
	10 mF to 100 mF	-	55 + 490000	-	-		
The expanded uncertainties given in this table expressed in mF/F + pF							

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Signatories:

1. Lam Sik Lee
2. Mohd Rahimi Bin Samsuddin (* EOS: Time base only, ** Excluded for EOS)

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Matrix G1 - AC Voltage Measurement

Range	FREQUENCY							
	(10 to 20) Hz	(20 to 40) Hz	(0.04 to 20) kHz	(20 to 50) kHz	(50 to 100) kHz	(100 to 300) kHz	(300 to 500) MHz	(0.5 to 1.0) MHz
0.22 mV to 2.2 mV	0.62 + 0.0051	0.24 + 0.0051	0.12 + 0.0051	0.42 + 0.0051	0.96 + 0.0079	1.2 + 0.015	1.9 + 0.028	3.9 + 0.028
2.2 mV to 22 mV	0.62 + 0.0057	0.24 + 0.0057	0.12 + 0.0057	0.42 + 0.0057	0.96 + 0.0079	1.2 + 0.014	1.9 + 0.028	3.9 + 0.028
22 mV to 220 mV	0.62 + 0.015	0.24 + 0.0091	0.12 + 0.0091	0.36 + 0.0091	0.96 + 0.028	1.2 + 0.028	1.9 + 0.040	3.9 + 0.12
0.22 V to 2.2 V	0.57 + 0.091	0.18 + 0.028	0.085 + 0.0068	0.14 + 0.018	0.28 + 0.079	0.49 + 0.15	1.2 + 0.040	2.5 + 0.96
2.2 V to 22 V	0.57 + 0.91	0.18 + 0.28	0.085 + 0.068	0.14 + 0.18	0.28 + 0.040	0.57 + 1.7	1.4 + 4.9	3.1 + 9.6
22 V to 220 V	0.57 + 9.1	0.18 + 2.8	0.091 + 0.91	0.25 + 4.0	0.57 + 9.1	1.7 + 100	5.3 + 100	13 + 220

The expanded uncertainties given in this table are expressed in mV/V + mV

Matrix G2 - AC Voltage Measurement

Range	FREQUENCY					
	(15 to 50) Hz	(0.05 to 1) Hz	(1 to 3) kHz	(3 to 10) kHz	(10 to 20) kHz	(20 to 30) kHz
220 V to 1100 V	0.45 + 18	0.091 + 4.0	-	-	-	-
220 V to 320 V	-	-	0.91 + 22	0.91 + 36	1.4 + 48	1.7 + 73
320 V to 800 V	-	-	0.91 + 71	0.91 + 120	1.4 + 180	1.7 + 240
800 V to 1050 V	-	-	0.90 + 150	0.91 + 240	1.4 + 360	-

The expanded uncertainties given in this table are expressed in mV/V + mV

Matrix H - AC Current Measurement

Range	FREQUENCY			
	(0.01 to 3) kHz	(3 to 10) kHz	(10 to 20) kHz	(20 to 30) kHz
0 μ A to 32 μ A	0.79 + 1	1.1 + 2	2.3 + 6.8	2.8 + 10
32 μ A to 320 μ A	0.79 + 0.35	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
0.32 mA to 3.2 mA	0.79 + 0.48	1.1 + 0.68	2.3 + 2.3	2.8 + 3.4
3.2 mA to 32 mA	0.79 + 4.7	1.1 + 7.3	2.3 + 15	2.8 + 25
32 mA to 320 mA	0.91 + 36	1.1 + 55	2.3 + 73	2.8 + 110
0.32 A to 3.2 A	1.1 + 570	2.8 + 2900	-	-
3.2 A to 10.5 A	2.3 + 3400	5.7 + 11000	-	-

The expanded uncertainties given in this table are expressed in mA/A + μ A

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 26 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Table I: DC Power

Voltage	Current	DC Power	CMC
329.9999 mV	3.29999 mA	(0.1089 to 1.089) mW	0.026 + 0.00056
	32.9999 mA	(1.089 to 10.89) mW	0.11 + 0.00043
	329.999 mA	(10.89 to 108.9) mW	0.11 + 0.0043
	2.19999 A	(0.0726 to 0.726) W	0.33 + 0.036
	11 A	(1.1063 to 3.63) W	0.65 + 0.29
3.299999 V	3.29999 mA	(1.089 to 10.89) mW	0.14 + 0.00041
	32.9999 mA	(10.89 to 108.9) mW	0.1 + 0.0043
	329.999 mA	(0.1089 to 1.089) W	0.1 + 0.044
	2.19999 A	(0.72 to 7.26) W	0.32 + 0.35
	11 A	(11.06 to 36.3) W	0.65 + 2.9
32.99999 V	3.29999 mA	(10.89 to 108.9) mW	0.14 + 0.0041
	32.9999 mA	(0.1089 to 1.089) W	0.1 + 0.043
	329.999 mA	(1.089 to 10.89) W	0.1 + 0.44
	2.19999 A	(7.26 to 72.6) W	0.32 + 3.5
	11 A	(0.1106 to 0.363) kW	0.65 + 29
329.9999 V	3.29999 mA	(0.1089 to 1.089) W	0.14 + 0.041
	32.9999 mA	(1.089 to 10.89) W	0.1 + 0.43
	329.999 mA	(10.89 to 108.9) W	0.11 + 4.3
	2.19999 A	(0.0726 to 0.726) kW	0.32 + 35
	11 A	(1.106 to 3.63) kW	0.65 + 290
1000 V	3.29999 mA	(0.33 to 3.3) W	0.08 + 0.53
	32.9999 mA	(3.3 to 33) W	0.05 + 5.4
	329.999 mA	(0.033 to 0.330) kW	0.06 + 54
	2.19999 A	(0.22 to 2.2) kW	0.21 + 500
	11 A	(3.35 to 11) kW	0.69 + 330

The expanded uncertainties given in this table expressed in mW/W + mW

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 27 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Table J: AC Power

Voltage	Current	AC Power	(10 to 20) Hz	(20 to 45) Hz	(45 to 65) Hz	(65 to 500) Hz	(0.5 to 1) kHz	(1 to 5) kHz	(5 to 10) kHz		
			The expanded uncertainties given in below table expressed in mW/W + uW								
32.999 mV	3.2999 mA	(0.0109 to 0.01089) mW	1.4 + 0.77	1.5 + 0.76	0.37 + 0.78	0.37 + 0.78	0.40 + 0.78	2.5 + 0.75	12 + 0.61		
	32.999 mA	(0.1089 to 1.0889) mW	4.2 + 0.52	3.8 + 0.54	1.9 + 0.65	1.9 + 0.65	1.9 + 0.64	6.0 + 0.38	17 + 0.17		
	329.99 mA	(1.089 to 10.889) mW	4.5 + 2.2	4.1 + 2.3	2.2 + 2.3	2.2 + 2.3	2.3 + 2.3	6.2 + 1.0	17 + 0.48		
	2.19999 A	(7.260 to 72.597) mW	4.5 + 16	4.5 + 16	2.2 + 16	2.2 + 16	2.3 + 16	9.6 + 6.6	2.1 + 93		
	11 A	(36.299 to 362.989) mW	-	-	2.1 + 93	2.2 + 91	4.0 + 70	-	-		
329.999 mV	3.2999 mA	(0.109 to 1.089) mW	3.0 + 0.48	2.5 + 0.52	0.90 + 0.46	0.77 + 0.68	0.9 + 0.66	5.7 + 0.32	17 + 0.14		
	32.999 mA	(1.089 to 10.890) mW	3.4 + 0.69	2.9 + 0.72	1.1 + 0.58	1.1 + 0.58	1.3 + 0.52	5.9 + 0.17	17 + 0.12		
	329.99 mA	(10.89 to 108.90) mW	3.4 + 7.1	2.9 + 7.4	1.1 + 4.6	1.1 + 6.3	1.3 + 5.8	5.9 + 1.1	17 + 1.3		
	2.19999 A	(72.599 to 725.99) mW	4.1 + 0.23	4.1 + 0.23	1.2 + 60	1.2 + 60	1.4 + 55	9.5 + 26	-		
	11 A	(0.3630 to 3.63) W	-	-	0.21 + 7800	0.38 + 7700	2.2 + 6700	-	-		
3.29999 V	3.2999 mA	(1.0890 to 10.890) mW	2.6 + 0.44	1.9 + 0.47	1.1 + 0.42	1.2 + 0.41	1.3 + 0.37	5.9 + 0.13	17 + 0.11		
	32.999 mA	(10.890 to 108.90) mW	2.6 + 4.4	1.9 + 4.7	1.1 + 4.3	1.1 + 5.3	1.2 + 3.8	5.9 + 1.3	17 + 1.1		
	329.99 mA	(0.1089 to 1.089) W	2.6 + 58	1.9 + 52	1.0 + 52	1.0 + 52	1.2 + 47	5.9 + 6.8	17 + 12		
	2.19999 A	(0.7260 to 7.260) W	2.7 + 420	2.7 + 420	1.9 + 440	1.2 + 460	1.3 + 420	9.4 + 240	-		
	11 A	(3.630 to 36.30) W	-	-	0.71 + 5900	1.1 + 5100	3.5 + 3200	-	-		
32.9999 V	3.2999 mA	(10.89 to 108.90) mW	2.6 + 4.8	1.9 + 5.2	1.2 + 5.0	1.6 + 3.9	1.3 + 4.5	5.9 + 1.5	17 + 1.2		
	32.999 mA	(0.1089 to 1.089) W	2.6 + 44	1.9 + 47	1.1 + 42	1.1 + 42	1.2 + 38	5.9 + 13	17 + 11		
	329.99 mA	(1.089 to 10.890) W	2.6 + 410	1.9 + 430	1.1 + 350	1.1 + 350	1.3 + 310	5.9 + 34	17 + 100		
	2.19999 A	(7.260 to 72.599) W	2.7 + 4200	2.7 + 4200	1.2 + 4600	1.2 + 4600	1.3 + 4200	9.4 + 2400	-		
	11 A	(36.30 to 363.0) W	-	-	0.75 + 58000	1.2 + 51000	3.5 + 32000	-	-		
			The expanded uncertainties given in below table expressed in mW/W + mW								
329.999 V	3.2999 mA	0.1089 W to 1.089 W	-	-	1.2 + 0.041	1.2 + 0.041	1.4 + 0.037	6.0 + 0.015	17 + 0.012		
	32.999 mA	1.089 W to 10.89 W	-	-	0.94 + 2.1	0.94 + 2.1	1.1 + 2.0	5.9 + 0.53	17 + 0.25		
	329.99 mA	10.89 W to 108.9 W	-	-	1.1 + 4.2	1.1 + 4.1	1.3 + 3.7	6.0 + 1.5	17 + 1.2		
	2.19999 A	72.60 W to 726.0 W	-	-	1.2 + 45	1.3 + 45	1.4 + 42	9.4 + 24	-		
	11 A	363.0 W to 3630.0 W	-	-	0.81 + 560	1.2 + 500	3.5 + 320	-	-		
1000 V	3.2999 mA	(0.330 to 3.30) W	-	-	1.1 + 0.59	1.1 + 0.59	1.3 + 0.56	6.2 + 0.2	17 + 0.19		
	32.999 mA	(3.30 to 33.0) W	-	-	1.2 + 1.2	0.99 + 6.1	1.2 + 5.8	6.2 + 2.0	17 + 1.9		
	329.99 mA	(33.0 to 330.0) W	-	-	1.1 + 51	1.1 + 51	1.2 + 47	6.2 + 17	17 + 17		
	2.19999 A	(220.0 to 2200.0) W	-	-	1.1 + 560	1.1 + 560	1.2 + 540	9.6 + 180	-		
	11 A	(1100.0 to 11000.0) W	-	-	0.92 + 730	1.3 + 700	3.6 + 590	-	-		

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 28 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

Matrix K - AC Voltage Source

Range	FREQUENCY							
	(10 – 40) Hz	(40 – 100) Hz	(0.1 – 2) kHz	(2 – 10) kHz	(10 – 30) kHz	(30 – 100) kHz	(100 – 300) kHz	(0.3 – 1) MHz
0 to 200 mV	0.27 + 0.0044	0.25 + 0.0045	0.23 + 0.0023	0.23 + 0.0045	0.48 + 0.0089	0.90 + 0.022	-	-
0.2 V to 2 V	0.20 + 0.022	0.17 + 0.022	0.15 + 0.022	0.17 + 0.022	0.28 + 0.045	0.57 + 0.23	3.4 + 2.3	11 + 23
2 V to 20 V	0.20 + 0.22	0.17 + 0.22	0.15 + 0.22	0.17 + 0.22	0.28 + 0.45	0.57 + 2.3	3.4 + 2.3	11 + 230
20 V to 200 V	0.20 + 2.2	0.17 + 2.2	0.15 + 2.2	0.17 + 2.2	0.28 + 4.5	0.57 + 23	3.4 + 230	11 + 2300
200 V to 1000 V	0.22 + 10	0.20 + 10	0.20 + 10	0.20 + 10	0.30 + 22	0.60 + 110	-	-

The expanded uncertainties given in this table are expressed in mV/V + mV

Matrix L - AC Current Source

Range	FREQUENCY				
	(10 – 40) Hz	(40 – 300) Hz	(0.3 – 1) kHz	(1 – 5) kHz	(5 – 10) kHz
0 to 200 μ A	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	0.38 + 0.022	-
0.2 mA to 2 mA	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	0.36 + 0.22	-
2 mA to 20 mA	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	0.35 + 2.3	-
20 mA to 200 mA	0.36 + 22	0.36 + 22	0.36 + 22	0.36 + 22	-
0.2 A to 2 A	0.69 + 450	0.69 + 450	0.69 + 450	2.3 + 910	-
2 A to 10 A	0.63 + 110	0.61 + 100	0.61 + 48	0.61 + 100	0.61 + 100

The expanded uncertainties given in this table are expressed in mA/A + μ A

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 29 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: TEMPERATURE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Temperature Indicator / Recorder / Controller			
Type K	-250 °C to -200 °C -200 °C to -100 °C -100 °C to +100 °C +100 °C to +600 °C +600 °C to +1372 °C	0.66 °C 0.32 °C 0.23 °C 0.28 °C 0.32 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type J	-210 °C to -100 °C -100 °C to +800 °C +800 °C to +1000 °C +1000 °C to +1200 °C	0.30 °C 0.23 °C 0.26 °C 0.28 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type T	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 0 °C 0 °C to +400 °C	0.68 °C 0.32 °C 0.27 °C 0.21 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type E	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 100 °C 100 °C to 1000 °C	0.66 °C 0.26 °C 0.21 °C 0.25 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type B	500 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1400 °C 1400 °C to 1820 °C	0.63 °C 0.47 °C 0.39 °C 0.43 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type N	-200 °C to -100 °C -100 °C to 900 °C 900 °C to 1100 °C 1100 °C to 1300 °C	0.67 °C 0.32 °C 0.26 °C 0.21 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: TEMPERATURE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Temperature Indicator / Recorder / Controller			
PT100 (PT385 & PT392)	-200 °C to -100 °C -100 °C to 100 °C 100 °C to 630 °C 630 °C to 850 °C	0.18 °C 0.13 °C 0.23 °C 0.34 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
2. Temperature Sensors			
THERMOCOUPLE / TEMPERATURE PROBE	0.0 °C 20 °C 20 °C to 50 °C 50 °C to 200 °C 200 °C to 420 °C 420 °C to 550 °C 550 °C to 660 °C	0.074 °C 0.08 °C 0.085 °C 0.30 °C 0.31 °C 0.35 °C 0.37 °C	Comparison with Dry Block RTD Probe & ICE Point PRT Probe
PT100	0.01 °C 20 °C to 50 °C 50 °C to 100 °C 100 °C to 200 °C 200 °C to 420 °C	0.0057 °C 0.035 °C 0.10 °C 0.11 °C 0.14 °C	Comparison with Standard PRT Probe in calibration Triple Point of Water Cell & Dry Block
3. Temperature & Humidity Indicator			
THERMOHYGROGRAPH / THERMOHYGROMETER	20 °C to 40 °C 41 °C to 60 °C	0.68 °C 1.7 °C	Comparison with Reference Temperature & Humidity Meter in Temperature Chamber
	30 %RH to 95 %RH	2.4 %RH	Comparison with Reference Temperature & Humidity Meter in Temperature Chamber

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 122

SCOPE OF CALIBRATION: TEMPERATURE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
4. Temperature Calibrator / Simulator			
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.40 °C	By electrical measurement using Temperature Calibrator
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C	By electrical measurement using Temperature Calibrator
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.63 °C 0.24 °C 0.16 °C 0.14 °C	By electrical measurement using Temperature Calibrator
Type E	-250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C	0.50 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	By electrical measurement using Temperature Calibrator
Type B	600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1550 °C 1550 °C to 1820 °C	0.44 °C 0.34 °C 0.30 °C 0.33 °C	By electrical measurement using Temperature Calibrator
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.40 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C	By electrical measurement using Temperature Calibrator

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatories:**

1. Lam Sik Lee
2. Mohd Rahimi Bin Samsuddin (Except for item 2 & 4)

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 32 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: TEMPERATURE

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
5. Temperature Indicator / Recorder / Controller			
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to +120 °C +120 °C to +1000 °C +1000 °C to +1372 °C	0.38 °C 0.22 °C 0.19 °C 0.30 °C 0.46 °C	By electrical simulation using Temperature Calibrator
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to +150 °C +150 °C to +760 °C +760 °C to +1200 °C	0.32 °C 0.19 °C 0.17 °C 0.21 °C 0.27 °C	By electrical simulation using Temperature Calibrator
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to +120 °C +120 °C to +400 °C	0.72 °C 0.28 °C 0.19 °C 0.17 °C	By electrical simulation using Temperature Calibrator
Type E	-250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C	0.50 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type B	600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1550 °C 1550 °C to 1820 °C	0.44 °C 0.34 °C 0.30 °C 0.33 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.40 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 33 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: TEMPERATURE

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
5. Temperature Indicator / Recorder / Controller			
PT100 (PT385 & PT392)	-200 °C to -100 °C -100 °C to 100 °C 100 °C to 630 °C 630 °C to 850 °C	0.18 °C 0.13 °C 0.23 °C 0.34 °C	By electrical simulation using Temperature Calibrator and reference table to ITS-90
6. Temperature Sensor			
THERMOCOUPLE / TEMPERATURE PROBE	0.0 °C 50 °C to 200 °C 200 °C to 420 °C 420 °C to 550 °C 550 °C to 660 °C	0.074 °C 0.30 °C 0.31 °C 0.35 °C 0.37 °C	Comparison with Dry Block RTD Probe & ICE Point PRT Probe
7. Temperature Calibrator / Simulator			
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.40 °C	By Electrical measurement using Temperature Calibrator
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C	By Electrical measurement using Temperature Calibrator
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.63 °C 0.24 °C 0.16 °C 0.14 °C	By Electrical measurement using Temperature Calibrator

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Schedule

Issue date: 6 October 2017
Valid until: 21 October 2020



MS ISO/IEC 17025

Page: 34 of 34

NO: SAMM 122

SCOPE OF CALIBRATION: TEMPERATURE

SITE CALIBRATION: CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
7. Temperature Calibrator / Simulator			
Type E	-250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C	0.50 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	By Electrical measurement using Temperature Calibrator
Type B	600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1550 °C 1550 °C to 1820 °C	0.44 °C 0.34 °C 0.30 °C 0.33 °C	By Electrical measurement using Temperature Calibrator
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.40 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C	By Electrical measurement using Temperature Calibrator
8. Temperature Controlled Enclosure			
TEMPERATURE CONTROLLED ENCLOSURE	-20 °C to 100 °C 100 °C to 400 °C	1.9 °C 2.0 °C	Calibration using Thermocouple Sensor & meter (Based on G-20 Document Thailand)

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

Signatories:

1. Lam Sik Lee
2. Mohd Rahimi Bin Samsuddin (Except for item 6, item 7 & item 8)